

## **Neuroenhancement using non-invasive brain stimulation**

Andrea Antal<sup>1</sup>, Bruce Luber<sup>2</sup>, Anna-Katherine Brem<sup>3,4,28</sup>, Marom Bikson<sup>5</sup>, Andre R Brunoni<sup>6,7</sup>, Roy Cohen Kadosh<sup>8</sup>, Veljko Dubljevic<sup>9</sup>, Shirley Fecteau<sup>10</sup>, Florinda Ferreri<sup>11,12</sup>, Agnes Flöel<sup>13,14</sup>, Mark Hallett<sup>15</sup>, Roy Hamilton<sup>16</sup>, Christoph S Herrmann<sup>17</sup>, Michal Lavidor<sup>18</sup>, Collen Loo<sup>19</sup>, Caroline Lustenberger<sup>20</sup>, Sergio Machado<sup>21,22</sup>, Carlo Miniussi<sup>23</sup>, Vera Moliadze<sup>24</sup>, Michael A Nitsche<sup>25</sup>, Simone Rossi<sup>26</sup>, Paolo M. Rossini<sup>27</sup>, Emiliano Santarnecchi<sup>28</sup>, Margitta Seeck<sup>29</sup>, Gregor Thut<sup>30</sup>, Zsolt Turi<sup>31</sup>, Yoshikazu Ugawa<sup>32</sup>, Ganesan Venkatasubramanian<sup>33</sup>, Nicole Wenderoth<sup>20,34</sup>, Anna Wexler<sup>35</sup>, Ulf Ziemann<sup>36,37</sup>, Walter Paulus<sup>38</sup>

<sup>1</sup>Department of Neurology, University Medical Center, Göttingen, Germany

<sup>2</sup>Noninvasive Neuromodulation Unit, Experimental Therapeutics and Pathophysiology Branch, National Institute of Mental Health, Bethesda, Maryland, United States

<sup>3</sup>University Hospital of Old Age Psychiatry, University of Bern, Bern, Switzerland

<sup>4</sup>Department of Neuropsychology, Lucerne Psychiatry, Lucerne, Switzerland

<sup>5</sup>Biomedical Engineering at the City College of New York (CCNY) of the City University of New York (CUNY), NY, USA

<sup>6</sup>Departamento de Clínica Médica e de Psiquiatria, Faculdade de Medicina da Universidade de São Paulo, São Paulo, Brazil

<sup>7</sup>Service of Interdisciplinary Neuromodulation (SIN), Laboratory of Neurosciences (LIM-27), Institute of Psychiatry, Hospital das Clínicas da Faculdade de Medicina da USP, São Paulo, Brazil

<sup>8</sup>School of Psychology, Faculty of Health & Medical Sciences, University of Surrey, Guildford, GU2 7XH, UK

<sup>9</sup>Science, Technology and Society Program, College of Humanities and Social Sciences, North Carolina State University, Raleigh, NC, USA

<sup>10</sup>Department of Psychiatry and Neurosciences, Faculty of Medicine, Université Laval; CERVO Brain Research Centre, Centre intégré universitaire en santé et services sociaux de la Capitale-Nationale, Quebec City, Quebec, Canada

<sup>11</sup>Unit of Neurology, Unit of Clinical Neurophysiology, Study Center of Neurodegeneration (CESNE), Department of Neuroscience, University of Padua, Padua, Italy.

<sup>12</sup>Department of Clinical Neurophysiology, Kuopio University Hospital, University of Eastern Finland, Kuopio, Finland.

<sup>13</sup>Department of Neurology, Universitätsmedizin Greifswald, 17475 Greifswald, Germany

<sup>14</sup>German Centre for Neurodegenerative Diseases (DZNE) Standort Greifswald, 17489 Greifswald, Germany

<sup>15</sup>Human Motor Control Section, National Institute of Neurological Disorders and Stroke, National Institutes of Health, Bethesda, MD, USA

<sup>16</sup>Department of Neurology, University of Pennsylvania, Philadelphia, PA, USA

<sup>17</sup>Experimental Psychology Lab, Department of Psychology, Carl von Ossietzky Universität, Oldenburg, Germany

<sup>18</sup>Department of Psychology and the Gonda Brain Research Center, Bar Ilan University, Israel

<sup>19</sup>School of Psychiatry & Black Dog Institute, University of New South Wales; The George Institute; Sydney, Australia

<sup>20</sup>Neural Control of Movement Lab, Institute of Human Movement Sciences and Sport, Department of Health Sciences and Technology, ETH Zurich, 8092 Zurich, Switzerland

<sup>21</sup>Department of Sports Methods and Techniques, Federal University of Santa Maria, Santa Maria, Brazil

<sup>22</sup>Laboratory of Physical Activity Neuroscience, Neurodiversity Institute, Queimados-RJ, Brazil

<sup>23</sup>Center for Mind/Brain Sciences – CIMeC & Centre for Medical Sciences - CISMed, University of Trento, Rovereto, Italy

<sup>24</sup>Institute of Medical Psychology and Medical Sociology, University Medical Center Schleswig Holstein, Kiel University. Kiel, Germany

<sup>25</sup>Department Psychology and Neurosciences, Leibniz Research Centre for Working Environment and Human Factors at TU Dortmund, Germany; Dept. Neurology, University Medical Hospital Bergmannsheil, Bochum, Germany

<sup>26</sup>Siena Brain Investigation & Neuromodulation Lab (Si-BIN Lab), Unit of Neurology and Clinical Neurophysiology, Department of Medicine, surgery and Neuroscience, University of Siena, Italy

<sup>27</sup>Department of Neuroscience & Neurorehabilitation, Brain Connectivity Lab, IRCCS-San Raffaele-Pisana, Rome, Italy

<sup>28</sup>Berenson-Allen Center for Noninvasive Brain Stimulation, Department of Neurology, Beth Israel Deaconess Medical Center, Harvard Medical School, Boston, MA, United States

<sup>29</sup>Département des Neurosciences Cliniques, Hôpitaux Universitaires de Genève, Switzerland

<sup>30</sup>Centre for Cognitive Neuroimaging, School of Psychology and Neuroscience, University of Glasgow, UK

<sup>31</sup>Department of Neuroanatomy, Institute of Anatomy and Cell Biology, Faculty of Medicine, University of Freiburg, Freiburg, Germany

<sup>32</sup>Department of Human Neurophysiology, Fukushima Medical University, Fukushima, Japan

<sup>33</sup>Department of Psychiatry, National Institute of Mental Health and Neurosciences, Bengaluru, India

<sup>34</sup>Future Health Technologies, Singapore-ETH Centre, Campus for Research Excellence And Technological Enterprise (CREATE), Singapore

<sup>35</sup>Department of Medical Ethics and Health Policy, University of Pennsylvania Perelman School of Medicine, Philadelphia, PA, USA

<sup>36</sup>Department of Neurology & Stroke, University of Tübingen, Germany

<sup>37</sup>Hertie Institute for Clinical Brain Research, University of Tübingen, Germany

<sup>38</sup>Department of Neurology, Ludwig Maximilians University Munich, Germany

### **Financial support:**

ML and AA are supported by the Volkswagen Foundation German 2018 - Israeli Cooperation in Biological and Life Sciences, Medicine under Grant [number A128416]. AA is additionally supported by the BMBF (Stimcode, 01FP2124B).

AF is supported by the Deutsche Forschungsgemeinschaft (327654276 – SFB 1315 to AF).

BL is supported by the NIMH Intramural Research Program (ZIAMH002955).

VM acknowledges the financial support by STIPED funding from the European Union's Horizon 2020 research and innovation programme (no. 731827).

G Venkatasubramanian acknowledges the support of Department of Biotechnology (DBT) - Wellcome Trust India Alliance (IA/CRC/19/1/610005) and Department of Biotechnology, Government of India (BT/HRD-NBA-NWB/38/2019-20(6)).

ARB receives scholarships and support from FAPESP (18/21722-8, 19/06009-6), the Brazilian National Council of Scientific Development (CNPq-1B), University of São Paulo Medical School (PIPA - FMUSP), and the UK Academy of Medical Sciences (Newton Advanced Fellowship), and the International Health Cohort Consortium (IHCC). ARB has equity in Flow Neuroscience, and receives in-kind support from Soterix Medical, MagVenture, Flow Neuroscience, Neurosoft, and NeuroConn.

The City University of New York holds patents on brain stimulation with MB as inventor. MB has equity in Soterix Medical Inc. MB consults, received grants, assigned inventions, and/or serves on the SAB of SafeToddlers, Boston Scientific, GlaxoSmithKline, Biovisics, Mecta, Lumenis, Halo Neuroscience, Google-X, i-Lumen, Humm, Allergan (Abbvie).

CSH holds a patent on brain stimulation and was funded by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) under Germany's Excellence Strategy – EXC 2177/1 – Project ID 390895286.

CL is supported by the Swiss National Science Foundation (PZ00P3\_179795).

C Loo and her research program receive competitive grant funding from the Australian National Health and Medical Research Council.

FF and her research program are partially supported by the Italian National Institute of Health, Grant number GR-2016-02361802.

MH is supported by the NINDS Intramural Research Program. He is an inventor of patents held by NIH for an immunotoxin for the treatment of focal movement disorders and the H-coil for magnetic stimulation; in relation to the latter, he has received license fee payments from the NIH (from Brainsway). He is on the Medical Advisory Boards of CALA Health and Brainsway (both unpaid positions). He is on the Editorial Board of approximately 15 journals and receives royalties and/or honoraria from publishing from Cambridge University Press, Oxford University Press, Springer, Wiley, Wolters Kluwer, and Elsevier. He has research grants from Medtronic, Inc. for a study of DBS for dystonia and CALA Health for studies of a device to suppress tremor.

MAN is supported by Deutsche Forschungsgemeinschaft (DFG) - Projektnummer 316803389 – SFB 1280 "Funded by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation) - Projektnummer 316803389 - SFB 1280, the German Federal Ministry of Education and Research (BMBF, GCBS grant 01EE1501), and the EU (Neurotwin, H2020, grant 101017716).

NW is supported by the Swiss National Science Foundation (320030\_175616) and the National Research Foundation, Prime Minister's Office, Singapore under its Campus for Research Excellence and Technological Enterprise (CREATE) programme (FHT).

YU is supported by research grants from the Research Project Grant-in-aid for Scientific Research from the Ministry of Education, Culture, Sports, Science and Technology (grants number: 18K10821, 10H01091, 20K07866).

UZ received grants from the German Ministry of Education and Research (BMBF), European Research Council (ERC), German Research Foundation (DFG), Janssen Pharmaceuticals NV and Takeda Pharmaceutical Company Ltd., and consulting fees from Bayer Vital GmbH, Pfizer GmbH and CorTec GmbH, all not related to this work.

RH is currently supported by research grants from the NIH (R01AG059763) and U.S. Department of Defense (PR191513). He has received fees as a speaker by Starfish Neuroscience and Alexion Pharmaceuticals. None of these sources of support are related to this work.

WP is currently supported by a research grant of the EU Joint Programme – Neurodegenerative Disease Research (*JPND*)

## **Abstract**

Attempts to enhance human memory and learning ability have a long tradition in science. This topic has recently gained substantial attention because of the increasing percentage of older individuals worldwide and the predicted rise of age-associated cognitive decline in brain functions. Transcranial brain stimulation methods, such as transcranial magnetic (TMS) and transcranial electric (tES) stimulation, have been extensively used in an effort to improve cognitive functions in humans.

Here we summarize the available data on low-intensity tES for this purpose, in comparison to repetitive TMS and some pharmacological agents, such as caffeine and nicotine. There is no single area in the brain stimulation field in which only positive outcomes have been reported. For self-directed tES devices, how to restrict variability with regard to efficacy is an essential aspect of device design and function. As with any technique, reproducible outcomes depend on the equipment and how well this is matched to the experience and skill of the operator. For self-administered non-invasive brain stimulation, this requires device designs that rigorously incorporate human operator factors. The wide parameter space of non-invasive brain stimulation, including dose (e.g., duration, intensity (current density), number of repetitions), inclusion/exclusion (e.g., subject's age), and homeostatic effects, administration of tasks before and during stimulation, and, most importantly, placebo or nocebo effects, have to be taken into account. The outcomes of stimulation are expected to depend on these parameters and should be strictly controlled. The consensus among experts is that low-intensity tES is safe so long as tested and accepted protocols (including, for example, dose, inclusion/exclusion) are followed and devices are used which follow established engineering risk-management procedures. Devices and protocols that allow stimulation outside these parameters cannot claim to be "safe" where they are applying stimulation beyond that examined in published studies that also investigated potential side effects.

Brain stimulation devices marketed for consumer use are distinct from medical devices because they do not make medical claims and are therefore not necessarily subject to the same level of regulation as medical devices (i.e., by government agencies tasked with regulating medical devices). Manufacturers must follow ethical and best practices in marketing tES stimulators, including not misleading users by referencing effects from human trials using devices and protocols not similar to theirs.

**Keywords:** neuroenhancement, cognitive enhancement, transcranial brain stimulation, tDCS, tACS, home-stimulation, DIY stimulation

### **Abbreviations and acronyms**

AD	Alzheimer's Disease
BDNF	Brain derived neurotrophic factor
DARPA	Defense Advanced Research Projects Agency
DIY	Do-It-Yourself
DLPFC	Dorsolateral prefrontal cortex
EEG	Electroencephalography
EMG	Electromyography
FCC	Federal Communications Commission -
FDA	(U.S.) Food and Drug Administration
IFCN	International Federation of Clinical Neurophysiology
LTD	Long-term depression
LTP	Long-term potentiation
MCI	Mild cognitive impairment
MDD	Medical Device Directive
MDR	Medical Device Regulation
MEP	Motor evoked potential
MRI	Magnetic resonance imaging
NIBS	Noninvasive brain stimulation
OTC	Over-The-Counter
PAS	Paired associative stimulation
PET	Positron emission tomography
PPC	Posterior parietal cortex
QPS	Quadripulse stimulation
RMT	Resting motor threshold
rTMS	Repetitive transcranial magnetic stimulation

SAE	Serious adverse event
SMA	Supplementary motor cortex
tACS	Transcranial alternating current stimulation
TBS	Theta-burst stimulation
tDCS	Transcranial direct current stimulation
tES	Transcranial electric stimulation
TMS	Transcranial magnetic stimulation
tRNS	Transcranial random noise stimulation

## 1. Introduction

Non-invasive brain stimulation (NIBS) interventions, such as repetitive transcranial magnetic (rTMS) and electric (tES) stimulation, impact perception, cognition, mood, motor activities, and other brain functions, both in healthy humans and patients (Buch et al. , 2017, Ekhtiari et al. , 2019, Huang et al. , 2009, Lefaucheur et al. , 2020, Lefaucheur et al. , 2017, Reis et al. , 2008). NIBS can improve, but also worsen, performance in competitive environments, school evaluations, athletic competitions up to the level of the Olympics, and in musicians' performance (e.g., Coffman et al., 2014 ; Luber & Lisanby 2014). Particularly, long-term effects of NIBS when used in uncontrolled ways (e.g., at-home environments) have not been fully explored. Consequently, the available data on these contexts are less frequently replicated the further they are away from experimental settings, although the possible benefits of NIBS in such contexts would have a substantial relevance in everyday life.

Neuroenhancement is perceived by the population as a source of promise and concern (Bard et al., 2018). In otherwise healthy subjects, the term 'neuroenhancement' can be conceptualized as a variety of interventions and technologies aiming to improve human performance above usual standards, beyond what is considered 'physiologically normal' either within this range or above this range. Neuroenhancement covers a broad spectrum of interventions, ranging from simple beverages with high caffeine levels and legal drugs such as nicotine to sophisticated technologies, like NIBS. The latter is the key focus of this paper. A European survey of attitudes relating to science and technology reported that biotechnology interventions (e.g., NIBS) are consistently viewed more favorably when applied therapeutically in patients than for enhancement in healthy individuals (Gaskell et al 2011). Nevertheless, 56% of the survey participants (mainly researchers and industry partners) approved of research aimed at enhancing human performance, with stronger positive regard for electrical brain stimulation (tES) compared to drugs. In our previously published safety guideline, we used the definition "*Neuroenhancement can be defined as any augmentation of core information processing systems in the brain apart from natural training, including the mechanisms underlying perception, attention, conceptualization, memory, reasoning, and motor performance*" (Antal et al. , 2017). A broader definition of human enhancement is put forth by the European SIENNA project (<https://zenodo.org/record/4066557#.YFsC1nkxmUn>): "*Human enhancement is a*

*modification aimed at improving human performance and brought about by science-based and/or technology-based interventions in or on the human body*” in various subfields: cognitive (mainly related to intelligence and memory); physical (sport); affective and emotional (well-being); moral (e.g., correcting decision-making behaviour, social norm compliance); cosmetic; longevity (anti-aging), and others. For instance, the potential of NIBS as a cognitive enhancer is described as “...interventions that improve cognitive abilities, including neuro-stimulation & neuromodulatory techniques,... These may impact personal identity, for instance, by altering someone’s moods, cognition, behaviour, and basic personality traits.” Based on other definitions, the subfields in that document are methodologically defined as “by reversibility, by relation to the body, by relation to therapy and by field or technique.” Since ‘enhancement’ aims to improve human performance beyond what is necessary to sustain or restore good health (Juengst and Eric, 1998), a difficult border definition between pathology and normality is needed, which in turn depends on social and cultural substrates and age. Here, the definition of enhancement can also be different when considering high-performance groups, such as professional athletes, musicians, or other professions that rely heavily on high-end performance. Nevertheless, it should also be stated that in the NIBS field only a small number of studies were designed for ‘meaningful neuroenhancement’, i.e. systematically developing protocols (e.g. which would require titration of the dose of stimulation) that result in relevant effects.

Neuroenhancement by NIBS is often the aim of at-home applications: This ideally should encompass a verified and safe protocol using a stimulator that has been certified according to country-specific (medical product) legal requirements, (e.g., <https://www.regulatory-affairs.org/en/development-excellence/news-page/verification-and-validation-of-medical-devices>), however, these requirements are not necessarily being met. Device manufacturers or social media contributors have frequently used scientific results from prior studies to back up and promote their devices or techniques (see (Santaracchi et al. , 2013). Social media users (e.g. <http://www.reddit.com>) recommend the application of low intensity tES for several purposes, including neuroenhancement, with no or only insufficient differentiation between CE-approved devices and home-made devices consisting of a 9-volt battery with two wires attached to the poles. In 2016, the International Federation of Clinical Neurophysiology (IFCN) warned against the self-(home)application of tES (so-called do-it-yourself or DIY), especially with non-approved devices

([https://wfneurology.org/news\\_events/archived-news/2016-01-25-ifcn](https://wfneurology.org/news_events/archived-news/2016-01-25-ifcn)). Since then, technical and methodological progress has expanded the diversity of possible stimulation techniques and related protocols. Here, we aim to provide a timely review of the present scientific knowledge and place it in context with neuronal (cognitive) enhancement by tES in otherwise healthy individuals, also including a comparison with rTMS and well-known neuroenhancing substances such as nicotine and caffeine. Different aspects of neuroenhancement, including possible long term effects, placebo effects, self-applied and -directed stimulation are discussed. Our main goal is to prevent uncritical and also mercantile biased referrals to the scientific framework available in the field.

## **2. Physical possibilities: rTMS and tES**

The two most frequently used non-invasive technologies for modulating cortical activity are repetitive transcranial magnetic stimulation (rTMS) and transcranial direct current stimulation (tDCS). For neuromodulation purposes, a single pulse of TMS needs to be delivered repetitively. During TMS, a strong ( $\sim 8000$  A) and brief ( $\sim 100$ - $300$   $\mu$ s) electric current passes through a coil inducing a rapidly changing magnetic field. This field change, unattenuated by the skull, induces, in turn, a secondary electric current in the brain. The induced currents can change neuronal activity according to the parameters of stimulation (Aberra et al. , 2020). The TMS-induced current intensity declines proportionally to the square root of distance; directly affecting cortical regions a few centimeters beneath the coil. More distant target neurons are influenced transsynaptically (Ardolino et al. , 2005, Lang et al., 2005). Simply due to much higher technical demands, combined with higher prices and safety concerns, rTMS presently plays a minor role in neuroenhancement in the domestic context. Its availability in research laboratories may be tempting for the researchers to apply to themselves, although a questionnaire revealed this to be the case in only 8 % of the researchers (Shirota et al. , 2014). In addition, the induction of seizures is rare, but still an issue in rTMS (Rossi et al. , 2021) but not in low-intensity tES (Antal et al., 2017). It is not surprising that in interviews of laypersons, 74% believed that *“It is essential that public authorities oversee and control neuro-enhancement”* (Bard et al., 2018).

Transcranial direct current stimulation (tDCS) delivers a low-intensity current to brain areas through two or more electrodes placed on the scalp. The neuromodulatory effect is

induced by the modulation of membrane polarization, which in turn, modulates spontaneous firing rates in a polarity-specific manner, based on direct current stimulation studies (Bindman et al. , 1964, Creutzfeldt et al. , 1962). In the cognitive domain it is important to differentiate between effects mediated by the target electrode and the return electrode. One option is to go for a small target electrode (e.g. 4 x 4 cm) warranting a higher current density and a large return electrode (e.g. 100 cm<sup>2</sup>) (Nitsche et al., 2007). Factors like electrode shape, size, connector position and conductivities of different electrode materials including saline solutions and electrode gels may influence the outcome (Saturnino 2015).

Aftereffects of both tDCS and rTMS are necessary to produce lasting behavioural changes, as subjects and patients cannot be stimulated continuously for prolonged periods. Aftereffects will likely build up better over a series of separate sessions (e.g., (Reis et al. , 2009)), a schedule that has been well-characterized in the treatment of depression. Repetitive NIBS sessions in human beings, spaced at intervals of several minutes (i.e., 3–30 min), have obtained greater and more durable changes in neuroplasticity responses than NIBS applied over more prolonged spacing periods, with the latter appearing to induce less stable, and rapidly reversible plasticity (Goldsworthy et al., 2015; Monte-Silva et al.,2013). It is assumed that long-term potentiation (LTP) and long-term depression (LTD) phenomena at the synaptic level are the morphological substrates for durable changes following NIBS (Liebetanz et al. , 2002; Nitsche et al. 2003, Rossini et al., 2019).

### **3. Comparative enhancement by tDCS, tACS and tRNS**

Transcranial alternating current (tACS)(Antal et al. , 2008) and random noise (tRNS) (Terney et al. , 2008) stimulation apply balanced alternating current not excluding paradigms with a DC offset (Marshall et al. , 2006). tACS may modulate the amplitude, frequency and phase of brain oscillations (Antal et al., 2008, Frohlich et al. , 2015, Helfrich et al. , 2014a, Helfrich et al. , 2014b) as does tRNS (mostly randomized frequencies between 100-640Hz)(Terney et al., 2008). These modulations can, in turn, alter cognitive functions (Cecere et al., 2015; Clayton et al., 2019, Fertoni et al. , 2011, Snowball et al. , 2013).

The weaker the stimulation intensity, the weaker the entrainment effect is on endogenous oscillations. With lower intensities also the frequency range allowing entrainment shrinks, following the so-called Arnold tongue principle (Voskuhl et al. , 2018).

tRNS may provide a less complex approach for the self-stimulation (home) user by circumventing the necessity to think about particular resonance frequencies to be targeted. The effect of a tRNS protocol depends on individual cognitive and neuronal factors as well as on age (Evans et al. , 2018, Frank et al. , 2018, Harty and Cohen Kadosh, 2019).

As with tDCS, both tACS and tRNS may fail to produce neural or behavioral effects (Brignani et al., 2013; Fekete et al., 2018; Lafon et al., 2017; Santarnecchi et al. , 2016, Schecklmann et al., 2021). A simple assumption is that stimulation intensity might have been too low to modulate neuronal activity in several studies (Antal et al., 2008, Lafon et al. , 2017, Voroslakos et al. , 2018). A more complex hypothesis assumes that increasing intensities cause increases in intracellular calcium increase with lower concentration ranges for achieving inhibition (LTD-like effects) or higher intensities achieving excitation (LTP-like effects (Moliadze et al., 2012, Shorafa et al., 2021)). In this case, a failure to induce inhibition or excitation might be caused by too low or too high intensity or an intensity in the transition zone between inhibition and excitation. Direct comparisons of the enhancement effects of tACS and tRNS have been inconclusive, although some have indicated that either one or both may be better than tDCS, depending on the type of the task (Berger et al. , 2019, Brem et al. , 2018, Fertoni et al., 2011, Inukai et al. , 2016, Mulquiney et al. , 2011, Murphy et al. , 2020). We believe that a general answer is not possible at present due to the variability of stimulation protocols, as well as of the hypotheses. Therefore, for the foreseeable future all three methods are worth further investigation.

#### **4. Considering cognitive/behavioral effects by NIBS: Theoretical, neurobiological and physiological framing of neuroenhancement**

Two studies in normal subjects have observed long-lasting effects on cognition in the order of months following a week of low-intensity tES (Snowball et al., 2013; Cappelletti et al. , 2013). The multiple-session mode naturally raises ethical questions regarding reversibility, concomitant cognitive adverse effects, and necessity of additional sessions for maintaining desirable effects (Huo et al. 2021).

Modifications of functional brain networks will underlie desired effects on cognition but these can be associated with changes in activity in other networks, either by a rebalancing of resource allocations or by direct (Brem et al., 2014) and also detrimental

interactions (Iuculano et al., 2013). Such trade-offs are likely under-recognized, as most NIBS studies can focus on only one or two tasks at a time, and may lead to detrimental and adverse effects. These may develop over time and only may become recognizable long after stimulation. For example, following a six-day cognitive training session, combined with anodal tDCS applied over the left dorsolateral prefrontal cortex (dlPFC) and cathodal over the right one, numerical learning was facilitated, but automaticity of performing the mathematical task was impaired when cathodal tDCS was applied over the posterior parietal cortex (PPC) (Iuculano and Cohen Kadosh, 2013). Therefore, it is not clear whether the reported negative effects represented a genuine cost of improving one function, or are explained through stimulation of the secondary brain site by the cathodal electrode, impairing its function and leading to poorer attention.

In a second example, one-Hz rTMS at the PPC increased detection accuracy when stimuli were presented in the field ipsilateral to the side of stimulation (Hilgetag et al., 2001) but decreased for stimuli presented in the visual field contralateral to the side of stimulation. A network approach based on a mathematical concept from game theory, i.e., the zero-sum game, attempts to explain this (Brem et al., 2014). The gains of one set of players at the game's end are matched by the losses of another set of players, resulting in a net sum of gains and losses of zero. Looking at the brain as a finite system with a limited processing capacity, the zero-sum game could be applied conceptually, such that neural resources applied towards one goal are taken from others. However, there is no clear demonstration of the zero-sum theory in the area of NIBS at present, therefore it is of more theoretical interest.

Integral to every function in the nervous system is a means to improve its own functioning. This can be achieved through conditioned learning and memory of even simple organisms that allow for prediction based on experience, through the top-down executive control in more advanced organisms that allows for learning to optimize the allocation of limited processing resources or through the automation of abilities that allows for smooth and effortless performance (compare the overwhelming experience of initially trying to drive a car and the almost unconscious execution of an experienced driver). When using NIBS to facilitate learning, the individual context must be taken into account. For example, there is no doubt that intensified training, e.g., in piano players or mathematicians, leads to top

performances with ceiling effects; tDCS here may only worsen performance (Furuya et al. , 2014, (Krause et al. , 2019). Overall, however, it will not come at a cost for a system to inhibit an irrelevant capacity when trying to perform a task. The “costs” come when stimulation is performed without carefully considering the dynamics and properties of the systems involved. Simultaneous NIBS/fMRI and NIBS/EEG, or investigation of state-dependence (Silvanto and Pascual-Leone, 2008) and endogenous oscillations (e.g., Zrenner et al., 2018) may identify network nodes and states through which NIBS could effectively enhance processing and others with a tendency towards deterioration. Specifically designed behavioral tests may then determine the “costs.”

Rebalancing abnormal activity levels between network nodes or reinforcing those underlying a task can be mechanisms for neuroenhancement (Lang et al., 2005; Sale et al., 2015). NIBS effects are state-dependent, being strongly influenced by the ongoing activity of the targeted region at the time of stimulation (Antal et al. , 2007) (Bortoletto et al. , 2015, Ferreri et al. , 2014, Fertonani et al. , 2014). Some of the effects of NIBS on network functions may be conveyed by interactions with brain oscillations (e.g. through local entrainment) as brain rhythms have been associated with communication between distant regions, depending on temporally correlated activity or interregional synchrony. In support of this, single or -better- repetitive TMS pulses have been shown to have the ability to phase-reset and align local oscillators in a given cortical region and to increase the amplitude of rhythms that such oscillators are most likely to generate (Paus et al., 2001; Rosanova et al., 2009; Thut et al., 2011). These effects have been considered to be potentially able to transiently modulate specific cognitive operations and behaviors (Ferreri et al 2014).

## **5. Comparing NIBS to pharmacological neuroenhancement, caffeine and nicotine**

Lifestyle-related substances, such as nicotine and caffeine, are associated with neuroenhancement. Data about the performance-enhancing effects of nicotine in healthy humans are relatively rare. Exam performance did not correlate with nicotine consumption, whereas exam-related anxiety did correlate positively (Kusturica et al. , 2019). The impact of nicotine on cognitive performance is critically affected by the habitual smoking state. In non-smokers, acute nicotine intake decreased executive functions (Grundey et al. , 2015) and also motor learning (Grundey et al. , 2017). Habitual smokers showed decreased executive and motor learning under nicotine deprivation compared with non-smokers. These cognitive

deficits were, however, reversed by exposure to nicotine (Grundey et al., 2015, Grundey et al., 2017). These results might mean that in healthy, non-smoking humans, nicotine probably has no neuroenhancing but instead performance-worsening effects, while chronic smokers require it for normal functioning. At the level of brain physiology, related effects of nicotine have been shown for tDCS-induced plasticity, and cortical excitability. Here, smokers showed reduced intracortical facilitation, and abolished LTP-like plasticity during nicotine withdrawal, effects that were reversed by nicotine consumption. In non-smokers, nicotine enhanced cortical inhibition, and converted LTP-like into LTD-like plasticity (Grundey et al. , 2013, Grundey et al. , 2012a, b). Smoking-state dependent effects of tDCS on cognitive performance have not been explored systematically, but might be relevant based on the impact of smoking state and nicotine on the physiological effects of tDCS.

Caffeine belongs to the methylxanthines and blocks adenosine receptors that mediate predominantly inhibitory effects. This disinhibition caused by adenosine leads to an increase in vigilance and attention (Daubner et al. , 2021). Studies in the field with pure caffeine during stimulation are rare. Most data on caffeine effects are based on energy drinks with confounding effects of other ingredients. With regard to mood, caffeine has been reported to have an increasing effect on anxiety, but not on other mood domains (Fiani et al. , 2021). With regard to cognition, caffeine was suggested to enhance processing speed, but not general attention (Fiani et al., 2021). The underlying motivations for caffeine intake are mainly concentration and memory enhancement and physical performance improvement (Cappelletti et al, 2015). Caffeine has a positive effect on sustained attention (Repantis et al., 2021). Interestingly, in the context with transcranial stimulation, caffeine mobilizes intracellular calcium (Cappelletti et al., 2015), which in turn plays a distinct role in determining either LTD- or LTP-like aftereffects of transcranial electric stimulation (Shorafa et al., 2021). Caffeine also appears to improve physical performance in both trained and untrained individuals (Guest et al., 2021).

Increased alertness as measured by pupillometry was reported after a 200 mg dose of caffeine (Zulkifly et al. , 2021b). In a direct comparison, tDCS was reported to be a more powerful fatigue countermeasure than caffeine (McIntire et al. , 2017). Conversely, caffeine intake only correlated with a practice effect on the grooved pegboard test in conjunction with 20 min tDCS, alone had no effect in this particular setup (Fagerlund et al. , 2015).

Caffeine has furthermore significant positive effects on both short- and long-term memory, albeit not in children (Fiani et al., 2021). Variability of plasticity effects in tES studies can be accentuated by uncontrolled caffeine consumption. Excitatory plastic aftereffects induced by 1 mA tACS at 140 Hz were turned into inhibition by espresso intake. Interestingly, decaffeinated espresso was not inert, since it reduced the respective stimulation-induced excitability enhancement to baseline. This effect may be due to residual, lower levels of caffeine in decaffeinated coffee or the up to 1,000 chemical components of coffee (Zulkifly et al. , 2020). The caffeine content in regular and decaffeinated coffees ranged from 10.9 mg/g to 16.5 mg/g and from 0.34 mg/g to 0.47 mg/g, respectively. In general accordance, 200 mg caffeine also reduced the excitatory effect of quadripulse stimulation with 5 ms intervals in intervention responders (Hanajima et al. , 2019).

The impact of caffeine on brain physiology is probably also dependent on habitual intake of the substance. In caffeine naïve subjects, 200 mg caffeine increased motor cortex excitability, and caffeine increased and prolonged the effects of LTP-like plasticity induction via paired associative stimulation with an interstimulus interval of 25 ms (PAS25)(Zulkifly et al. , 2021a), but caffeine did not alter 140 Hz tACS-induced excitability changes. In caffeine consumers, the effects were less clear. Interactions were shown between caffeine concentrations, baseline cortical excitability and cortisol levels (Zulkifly et al., 2021b).

In summary, the effects of pharmacological “neuroenhancers” such as nicotine and caffeine are confounding factors in brain stimulation protocols. Ideally, participants should not be mixed with regard to nicotine or caffeine consumption behavior. Simple deprivation of the respective substance likely does not solve the problem, because at least with respect to nicotine, withdrawal has a modulatory effect on stimulation outcomes itself.

## **6. Boosting intelligence and creativity, and the role of training in cognitive enhancement with tES**

Intelligence and creativity, cornerstones of civilization, have also been the target of NIBS research. Research to date has yielded heterogeneous results, likely deriving from using a range of different stimulation setups and varied quantification/translation of said concepts. “Intelligence” refers to a broad and complex cognitive concept that is associated with a range of life outcomes (Strenze, 2007) and is, therefore, an appealing target. A few

studies have combined NIBS with cognitive training and observed an improvement in fluid intelligence after applying tDCS and tRNS compared to tACS (Almquist et al., 2019; Brem et al., 2018). Administering 40 Hz tACS resulted in a shortened response latency when solving fluid intelligence tasks (Santarnecci et al., 2016) and an improvement, particularly in more difficult items (Neubauer et al., 2017; Pahor & Jausovec, 2014; Santarnecci et al., 2013) whereas brain stimulation alone either resulted in no improvement (40Hz, 10Hz or sham tACS, online) (Grabner et al., 2018) or even caused a detrimental effect (bilateral prefrontal or sham tDCS, offline) (Sellers et al., 2015).

Creativity can be defined as “*the use of imagination or production of original ideas to create something (useful)*” (Barron, 1955). According to this definition, tDCS modulated creativity as measured with tasks assessing self-focused attention and mind wandering, inhibitory control, and divergent and convergent thinking, i.e., creative thinking (Lucchiari et al., 2018). This line of research suggests that stimulation may be “unleashing” creativity, mainly by modulating the bilateral temporal and prefrontal cortices. Placing the cathode over left and the anode over right cortical areas, in accordance with the hemispheric balance hypothesis, led to several studies finding positive results (Chryssikou et al., 2021; Hertenstein et al., 2019; Ivancovsky et al., 2019; Lifshitz-Ben-Basat & Mashal, 2021; Mayseless & Shamay-Tsoory, 2015). According to this hypothesis, a shift in interhemispheric balance resulting in a predominance of the right over the left hemisphere promotes creativity. However, recent neuroimaging and lesion studies suggest increased creativity occurs after enhancing neuronal excitability in the left frontal areas associated mostly with verbal creativity, as measured with convergent and divergent thinking tasks, idea selection and reasoning (Cerrutti & Schlaug, 2009; Green et al., 2017); notwithstanding, nil findings in a verbal insight test after administering tDCS over the left and right anterior temporal lobe have been reported (Aihara et al., 2017). Enhanced creativity was also reported using bilateral prefrontal tRNS (Pena et al., 2019). Application of 10 Hz tACS over bilateral frontal areas was associated with improved figural creativity (Lustenberger et al., 2016) and marginally improved verbal creativity (Grabner et al., 2018), but no significant effects were found with 40 Hz tACS.

Pairing tES with cognitive and behavioral training regimens may be effective, when no consistent effect has been associated with stimulation alone (Cappelletti et al., 2013,

Mancuso et al., 2016), or cognitive training alone (Martin et al, 2014). The approach aligns well with some of the presumed neural mechanisms of tES, which involve modulating ongoing neural activation patterns by strengthening or weakening these patterns via Hebbian synaptic mechanisms of neuroplasticity (Jackson et al., 2016; Kronberg et al., 2020; Reato et al., 2013), in which patterns of ongoing neural activity are selectively modulated and reinforced by stimulation (Gill et al., 2015). This approach has also been employed in clinical samples to target cognitive functions, like language processing (Nissim et al., 2020), motor learning (Wang et al., 2021), attention (Boroda et al., 2020), working memory (Jones et al., 2015), multitasking ability (Filmer et al., 2017; Manor et al., 2016), numeracy (Looi et al., 2017, (Looi et al. , 2017) and error awareness (Harty et al., 2014).

One obstacle has been the high degree of heterogeneity of the tDCS parameters, including stimulation polarity, intensity, and duration, electrode placement, and the temporal relationship of stimulation to behavioral training tasks (Martin, Liu, Alonzo, Green, & Loo, 2014)(Ehrhardt et al., 2021; Weller et al., 2020). A second one is the identification of responders. Those whose performance was weaker at baseline may be more likely to experience performance gains than those whose baseline performance was stronger (Asseondi et al., 2021; Krebs et al., 2021; Sarkar et al., 2014; Turkeltaub et al., 2012). Finally, the degree to which training effects on a particular task during stimulation can generalize to related tasks is unclear (Andrews et al., 2011; Filmer et al., 2017, Looi et al. , 2016).

## **7. Neuroenhancement and motor learning**

With small magnitudes of effect, rTMS over the pre-supplementary motor cortex (preSMA) may enhance early motor learning processes, and over the supplementary motor cortex (SMA) may enhance the late motor learning process (execution speed-up)(Hean et al, 2020). Quadripulse stimulation (QPS) (Hamada et al, 2008, 2009) was found to be one of the most effective methods, as it induces long-term effects in most studies with low inter-individual variability (Tiksnadi et al, 2020) (Matsumoto et al, 2020). QPS may improve performance in a sequential learning task (Shimizu et al, 2020) as confirmed by human neuroimaging studies (Hikosaka et al, 1996; Sakai et al,1998). However, due to its technical requirements, QPS will likely not be broadly applicable for neuroenhancement in healthy persons. During tDCS over M1 implicit learning was speeded up compared with prefrontal

cortex or medial frontal cortex, and sham stimulation (Nitsche et al, 2003). However, as of 2021, very heterogeneous results emerge from the ~ 500 publications on motor learning and tDCS, and no clear recommendations can be made here (Buch et al. , 2017).

## **8. Neuroenhancement during sleep**

Sleep is a particularly interesting target for NIBS and enhancement because stimulation takes place in a state of unconsciousness that is vital to different brain and body processes (Grandner, 2017). One important function of sleep is memory consolidation, i.e., the reactivation and stabilization of fragile memory traces built during wakefulness (Rasch and Born, 2013). Accordingly, recall performance of declarative or non-declarative memories has been one of the most frequently employed paradigms to investigate whether modulating brain activity during sleep with tES elicits functional effects in healthy adults or those with a neurological disorder (Cellini and Mednick, 2019, Malkani and Zee, 2020). Brain oscillations that dominate non-rapid eye movement sleep (NREM) are thought to mediate the beneficial effect of sleep on various functions (Rasch and Born, 2013, Tononi and Cirelli, 2020). Non-invasive tES approaches that promote these oscillations have attracted tremendous interest in (i) establishing the functional role of sleep oscillations in brain and body processes, and (ii) restoring diminished functions in conditions that are associated with impaired sleep oscillations. In the laboratory environment, different tES applications ranging from oscillatory tDCS to sophisticated closed-loop tES systems that mimic the frequency of sleep oscillations have been mainly applied in healthy participants (e.g.(Cellini et al. , 2019, Jones et al. , 2018, Ketz et al. , 2018, Ladenbauer et al. , 2017, Ladenbauer et al. , 2021) but also in a few patient populations (Goder et al. , 2013, Ladenbauer et al., 2017, Lafon et al. , 2017, Prehn-Kristensen et al. , 2014). Due to pronounced tES artifacts in the EEG, most laboratory studies analyze EEG data during intermittent, stimulation-free intervals and additionally employ behavioral outcome measurements to infer the effectiveness of tES for modulating sleep and sleep-related processes (Malkani and Zee, 2020).

Most previous studies have applied bilateral frontal, slow oscillatory tDCS (SO-tDCS) protocols, typically at ~0.75 Hz mimicking the frequency of slow oscillations, to modulate sleep oscillations and sleep-related memory processes (reviewed by Grimaldi et al. , 2020). For this protocol, two anodal electrodes are placed over the left and right frontal cortex

(corresponding to the F3 and F4 locations) and the corresponding cathodes are placed over the left and right mastoid. The stimulation signal induces a current which oscillates between zero and the maximal current intensity that is typically chosen to give maximum current densities of up to 0.522mA/cm<sup>2</sup>. SO-tDCS is commonly applied for less than 30 min during the first period of stable NREM sleep, which is interleaved with stimulation-free, short intervals in order to analyze slow wave activity and sleep spindles. Less frequently used stimulation protocols employ (i) tACS around 1 Hz to entrain slow wave activity (Robinson et al. , 2018) or tACS at 12 Hz to entrain sleep spindles (Lustenberger et al. , 2016) and (ii) closed-loop setups whereby sleep oscillations are analyzed in real-time to trigger short bouts of SO-tDCS(Cellini et al., 2019) or 12-Hz tACS (Lustenberger et al., 2016).

The most consistent result across studies, which used SO-tDCS and tACS at slow wave frequencies, is increased slow wave activity in the stimulation-free windows compared with sham stimulation (e.g.(Cellini et al., 2019, Jones et al., 2018, Ketz et al., 2018, Ladenbauer et al., 2017, Ladenbauer et al., 2021). However, some studies were unable to reproduce this effect despite a similar montage (e.g.(Bueno-Lopez et al. , 2019, Eggert et al. , 2013, Koo et al. , 2018, Sahlem et al. , 2015) . By contrast, effects on spindle activity seem variable, as well as the effects on declarative and non-declarative memory tasks (for review see: Grimaldi et al., 2020). To date, stimulation parameters or inter-individual differences explaining the variability of effects on sleep and memory modulation have not been systematically investigated. However, it will be crucial to better understand these factors for a rational, individualized, and more successful design of sleep tES (Frohlich and Lustenberger, 2020, Koo et al., 2018).

There is an increasing interest in modulating sleep in unsupervised settings and ideally at home because (i) sleep quality might suffer through the laboratory environment, and (ii) laboratory settings are ill-suited for studying effects achievable via long-term neuromodulation during sleep. However, it is doubtful whether tES protocols as described above are suitable for neuromodulation outside of the laboratory. In particular, it is currently unclear whether a safe application can be enabled using tES considering that users are in an unconscious state and unintended side-effects might not directly be noted. Furthermore, tES during sleep is currently not available in an unsupervised setting because (i) it requires expert knowledge to correctly position the electrodes and reduce impedance,

and (ii) precise stimulation requires an artifact-free EEG and individualized online monitoring to target the correct sleep stages and oscillatory processes. Because possible long-term adverse effects of nightly tES are currently not known, and the functional effectiveness of modulating sleep with tES seems to be limited as evaluated by behavioral markers of memory consolidation, tES-based approaches have gained virtually no traction for sleep applications outside of the academic setting. On the other hand, wearables such as Apple Watch, Samsung Watch etc. provide some surveillance including pulse rate and oxygen saturation with increasing accuracy, which provide an initial step towards future developments for tES home-use.

## **9. Transcranial electrical stimulation for well-being, reducing stress and burnout symptoms**

Many companies marketing tDCS directly to use, indicate that devices are not sold for medical conditions but to enhance “wellness”, for example by decreasing general stress level, enhancing focus, or ameliorating burnout symptoms. When the psychology and behavioral sciences reporter, Miles O’Brien, underwent the procedure for a television interview, he said that his brain “seemed to turn on like a light bulb” and after, “It was like a jolt of caffeine without the tense feeling.” (<https://www.pbs.org/newshour/show/gentle-electrical-jolt-can-focus-sluggish-mind>). Given the strong placebo effects of NIBS, it is possible that sham stimulation would produce the same answer (see below in section 11). At “Wellness Clinics”, business leaders and diplomats are invited to come and relax with healthy food, spa treatments and brain stimulation. Although no controlled scientific trials related to ‘improving well-being’ have been published, the framework of scientific tES data is used to claim that a decrease of subjective and objective stress levels and burnout improvement can be achieved that generally can improve the quality of life.

The scientific framework for tES does include improvements in properties related to well-being. For example, burnout, described as a 'state of vital exhaustion' was added to the 10<sup>th</sup> edition of the International Statistical Classification of Diseases and Related Health Problems (ICD-10: Z73.0:WHO, 2011). Left DLPFC anodal tDCS coupled with right orbital cathodal stimulation (F8) applied for three weeks improved burnout symptoms related to attention (Van Noppen et al. , 2020). Chronic stress, e.g. due to the pandemic (Castelo-Branco and Fregni, 2020) may benefit from tES, as well as some cases of post-COVID

syndrome by tDCS at home stimulation (Eilam-Stock et al. , 2021). Anodal and cathodal tDCS over the left and right DLPFC, respectively, for a single 30 min session in individuals with high but not with low anxiety profiles (Sarkar et al. , 2014) improved performance in cognitive tests and was associated with lower cortisol levels than after sham tDCS. Three minutes of anodal and cathodal tDCS to the left DLPFC and right supraorbital area, respectively, significantly reduced ratings of unpleasantness in subjects exposed to distressing images, compared with sham stimulation (Boggio et al. , 2009). Petrocchi et al. (Petrocchi et al. , 2017) also observed an increase in positive affectivity in healthy subjects after 15 min of 2 mA tDCS over the left temporal lobe contrary to sham stimulation. Nevertheless, other studies found no mood-improving effects in healthy subjects (Morgan et al. , 2014, Motohashi et al. , 2013).

With regard to other possibilities, Ironside et al. (Ironside et al. , 2016) showed in healthy subjects that tDCS over the prefrontal cortex could reduce vigilance to threatening stimuli. Likewise, a similar treatment protocol improved frustration tolerance (Plewnia et al. , 2015). Brunoni et al. (Brunoni et al. , 2013) found that a single three-minute session of 1.5 mA anodal / cathodal tDCS targeting the left / right DLPFC led to an increase in high frequency heart rate variability and a decrease in salivary cortisol levels that was not seen with sham or cathodal stimulation. Stimulation might have an effect on the autonomic nervous system and the hypothalamic-pituitary-adrenal (HPA) axis, however in a recent study, PAS and tACS did not elicit changes in saliva corticosteroid concentrations (Zulkifly et al., 2021b).

In summary, research data provided preliminary evidence for the value of anodal tDCS over the left DLPFC in the subjective and objective reduction of stress. Therefore, repeated application of tDCS might prove to be a useful and affordable therapy option in addition to conventional therapy, or in combination with other holistic methods (e.g. meditation) to help patients with burnout and chronic stress. However, further research is needed in randomized controlled trials to substantiate the preliminary evidence. Moreover, at the present state of tES science, due the limited scientific evidence, the application of brain stimulation for 'wellness' in healthy subjects is not recommended.

## **10. Sport, brain stimulation, doping**

TDCS can be seen as a potential ergogenic resource to improve muscular strength (Lattari et al., 2016, 2020), and endurance (Okano et al., 2015; Lattari et al., 2018) in both nonathletes (Okano et al., 2015; Lattari et al., 2016, 2020; Machado et al., 2019; Angius et al., 2018) and athletes (Sales et al., 2016; Hazime et al., 2017; Vargas et al. 2018).

With regard to sport and NIBS, the most frequently targeted area is the left DLPFC, because it is responsible for the inhibition of motor areas. Generally, this area has a role in the modulation of plenty of functions and processes, such as emotion, mood, and memory. Therefore, it is difficult to measure the impact of a given stimulation on just one parameter. The left DLPFC is an area that acts in controlling fatigue and exertion, thus, it is one of the most important areas involved in deciding whether or not to continue exercising (Lattari et al., 2016, Tanaka et al., 2013). Physical fatigue is a complex phenomenon, and factors such as perceived effort and central inhibition may be involved. In fatigue, there is an increase in brain activity (beta power, measured by EEG) and a greater synchronization between the left and right DLPFC (Tanaka 2013). Fatigue may be compensated in the DLPFC when there is a decrease in motor cortex activation due to the presence of central fatigue induced by exercise (Menotti et al., 2014). In line with this, Lattari et al. (2020) found that anodal tDCS applied to left DLPFC and cathode over the Fp2 at 2 mA for 20 min improved volume load and perceived exertion compared with sham tDCS and with tDCS using a reversed montage for lower limbs exercise in healthy young individuals. Twenty minutes of anodal tDCS at 2 mA applied over the left DLPFC (cathode over the Fp2) increased the tolerance to the exercise performed in the cycloergometer with maximum load in moderately active women compared with sham stimulation (Lattari et al. 2018). This finding implies that the modification of the neuronal activity in the left DLPFC could enhance endurance exercise performance by maintaining the volitional impulse to the motor cortex. However, other studies have failed to show significant tDCS effects on exercise performance (Barwood et al. 2016; Kan et al. 2013; Muthalib et al. 2013; Angius et al. 2015). Early reports on facilitation of motor learning (Nitsche et al. , 2003a) were recently confirmed and extended to improvement of jump performance (Grosprêtre et al 2021). If data on skilled musicians could be transferred to sports, a generalized conclusion could be that the chance of improving gross motor power or endurance might be higher as compared to skilled sports requirements (Furuya et al., 2014).

Using tDCS to improve performance in sport/exercise could be considered a form of “neurodoping” (Davis 2013; Banissy and Muggleton 2013; Colzato et al. 2017; Park 2017), as it is associated with physical risks, behavioral and ethical issues. Nevertheless, tDCS has escaped so far the standards of the world anti-doping agency (WADA), which are entirely focused on drugs: [https://www.wada-ama.org/sites/default/files/resources/files/2021list\\_en.pdf](https://www.wada-ama.org/sites/default/files/resources/files/2021list_en.pdf). Since there is reasonable evidence that some tES protocols may improve motor performance in normal persons, it may thus be partially seen as a ‘doping method’, without legal regulations so far. Moreover, so far, there is no “biomarker” for detecting its use. It is suggested that tDCS should not be prohibited but WADA should adopt a similar approach to monitor its use as it currently does to caffeine (Pugh and Pugh, 2021). Prohibition would encourage the use of unsafe devices and protocols or stimulation. As part of that monitoring, WADA could work with stimulation companies developing tDCS devices to ensure that athletes are properly informed about the applications, risks and effects of tDCS. Furthermore, athletes should be supported to document and report the use of tDCS, and to encourage them to participate in controlled scientific studies.

## **11. Neuroenhancement and the placebo effect**

Developing even more effective sham tES procedures is crucial for assessing neuroenhancement by tES (Davis et al., 2014; Fonteneau et al., 2019; Greinacher et al., 2019; O’Connell et al., 2012; Turi et al., 2019). Although most tES studies use sham tES protocols, few have directly investigated placebo or nocebo effects (e.g., BinDawood et al., 2020; Petersen and Puthusserypady, 2019; Turi et al., 2017, 2018). Similar to other placebo interventions (Colagiuri et al., 2011), sham tES protocols may induce complex psychobiological responses despite having no direct physiological effects.

A placebo effect can be defined as an enhancement and a nocebo effect as an impairment of a given outcome variable during and/or after the application of sham tES. The outcome variable can refer to any biomarker and may extend to cognitive functions or electrophysiological responses (Petersen and Puthusserypady, 2019; Turi et al., 2018).

Both conscious or unconscious factors can contribute to the generation of placebo effects (Wager and Atlas, 2015). To begin with, most participants experience a moderate

amount of cutaneous sensation (e.g., itching, tingling) with conventional tES, which is supposed to be mirrored during sham tES (Fertonani et al., 2014; Matsumoto and Ugawa, 2017). For tACS, in addition to the cutaneous effects, visual flickering effects (i.e., phosphenes) may occur (Lorenz et al., 2019; Schutter and Hortensius, 2010; Turi et al., 2013). Blinding may be ineffective in tACS protocols with frequencies between ca. 5 and 40 Hz if no attention is paid to phosphenes (please cite Turi et al., 2013).

A commonly used sham tES method is the 'fade-in, short-stimulation, fade-out' (FSF) protocol (Ambrus et al., 2012), though it is important to note some forms of 'sham' that also include ongoing low intensity, microampere stimulation have been shown to have biologically active effects (Nikolin et al., 2018). Early tDCS studies suggested that participants were unable to distinguish between the real and FS or FSF protocols (Gandiga et al., 2006). However, subsequent studies provided accumulated evidence that the FSF protocols cannot completely blind the commonly used real tDCS protocols (Greinacher et al., 2019; O'Connell et al., 2012; Turi et al., 2019), in particular when intensities higher than 2 mA and stimulation durations longer than ten minutes are used.

Subjectively experiencing the tES-induced perceptual adverse effects can contribute to the generation of placebo or nocebo effects. Besides, a general limitation of single-blind studies is that they may inadvertently influence the interaction between the operator and the participant (O'Connell et al., 2012; Palm et al., 2013). For example, the operator might unintentionally give different instructions during the real than during the sham sessions. Instructional manipulations combined with sham tES could induce cognitive placebo or even nocebo effects in healthy individuals (Turi et al., 2018). Ineffective blinding may not only interfere with attentional processes during tES, but might lead to psychobiological responses through expectations (Colagiuri et al., 2011; Oken et al., 2008). Experimentally manipulating the expectancy of tES outcome, e.g., by placebo-inducing written instructions combined with sham tES can on itself influence cognitive functions in young healthy adults (Turi et al., 2017, 2018). The combination of sham tES with conditioning (i.e., surreptitiously manipulating the trial-by-trial feedback of the cognitive task) led to a placebo effect both in the subjectively reported cognitive effects, as well as in the instrumental learning performance (Turi et al., 2018). Placebo effects might be highest in children and lowest in patients with dementia

(Benedetti, Carlino, & Pollo, 2011; Gniss, Kappesser, & Hermann, 2020), which leads to the next section.

## 12. Neuroenhancement in paediatric population

Several randomized, placebo-controlled trials suggest beneficial effects of tDCS over cognitive functions in the treatment of dystonia, refractory epilepsy, attention deficit/hyperactivity disorder, and autism in children/adolescents (for reviews see: Andrade et al. , 2014, Doruk Camsari et al. , 2018, Finisguerra et al. , 2019, Lee et al. , 2019, Muszkat et al. , 2016, Palm et al. , 2016, Rajapakse and Kirton, 2013, Rivera-Urbina et al. , 2017, Salehinejad et al. , 2019). Children and adolescents differ from adults in the conductivity of their skull tissues, in white and gray matter content, and in CSF volume, and when compared to adults they also have a smaller brain-scalp distance, all of which influence the distribution and intensity of the electric field. For example, reducing the HD-tDCS 4x1 ring circumference from 5 cm to 2.5 cm for use on children's heads reduced peak electric field values to intensities broadly more comparable to 4x1 HD-tDCS with a 5 cm ring in adults (Kessler et al. , 2013). The age-related anatomical and physiological changes of the brain may even reverse stimulation effects (Moliadze et al. , 2018), Moliadze et al. , 2015b) at about half of the intensity which is known to reverse effects in adults (Batsikadze et al. , 2013b). Recent results support the notion of tDCS dosage administration based on subject-specific models (skull thickness, tissue electrical conductivities) in children (Hunold et al. , 2021).

Studies conducted in children have not reported more adverse effects than observed in adults (Andrade et al. , 2014; Moliadze et al. , 2015a, Moliadze et al. , 2015b, Splittgerber et al 2020 (Splittgerber et al., 2021). Special attention should be paid to muscle jerks on awakening by questioning adolescents and their parents in order to detect evidence of myoclonic epilepsy, which may indicate greater risks with tDCS (Sierawska et al., 2020; Splittgerber, et al., 2019). In terms of tolerability, there are no special concerns for using tES in pediatric patients. In terms of safety, limited data exists concerning long-term after-effects of tES in children/adolescents.

In summary, the risk in using tES for neuroenhancement in a pediatric population, especially as home-stimulation, has not been sufficiently studied. *“Children [cannot] be considered as ‘small adults’ when testing medical interventions”* (Davis, 2014), see also

(Maslen et al. , 2014). Therefore, the findings from studies in adults cannot simply be extended to studies in children since the evidence regarding adverse effects in pediatric patients is still limited. Researchers have to take into account that because of the neuroanatomical and physiological differences between the brains of children and adults, the tES parameters applied in adults will have a potentially greater impact when applied to the pediatric brain. Monitoring the occurrence of potential cumulative effects and possible impairments in the paediatric population is necessary. At present, tES in children for the purpose of neuroenhancement is discouraged.

### **13. Neuroenhancement or treatment? Neuroenhancement in the elderly and cognitively impaired population**

At the other end of the age spectrum, a growing elderly population has increased the prevalence of several neurodegenerative disorders like dementia due to Alzheimer's (AD) and mild cognitive impairment (MCI). "Normal" aging itself is associated with progressive decline in cognitive functions like memory, learning, and executive function (Prehn and Flöel, 2015). The direct and indirect costs due to "normal" cognitive deterioration not only cause a profound loss of quality of life for the individuals, but present significant challenges to the sustainability of the welfare system, an issue exacerbated in neurodegenerative disease (Colzato et al. 2020).

Given the dearth of pharmacological interventions for cognitive enhancement in the elderly population, non-pharmacological avenues include cognitive training, physical activity, nutraceuticals, and NIBS techniques, particularly tDCS and rTMS (Chu et al. 2021), have attracted increasing attention (Prehn and Flöel, 2015). Anodal tDCS tailored cognitive training with stimulation of the left DLPFC may be beneficial for several subdomains of cognitive dysfunction in patients with MCI or AD (Chu et al. 2021) in line with cognitive training with anodal stimulation of the right temporoparietal cortex, which may have positive effects on object location memory in patients with MCI (De Sousa et al, 2020).

The variability of NIBS effects in older adults may be greater due to aging-related brain atrophy. Furthermore, the increased intracranial CSF content leads to a broader current distribution. However, using predictive current flow models based on structural

brain images may help personalize the electrode positions with resulting improvement in target localization (Antonenko et al., 2021).

#### **14. Personalized neuroenhancement using cognitive and behavioral fingerprints: opportunities and challenges**

At present, cost and time requirements will prevent a general, broad application of neuroenhancement using NIBS techniques, mainly the application of rTMS. Intuitively, the general approach to precision medicine instead of a one-dose-fits-all mode can be translated to neuroenhancement applications in both healthy individuals and patients, and even more so when utilizing NIBS techniques and technology where individual differences in the response to brain stimulation protocols constitute the norm rather than the exception (Bikson et al., 2018; Corp et al., 2020; Santarnecchi et al., 2016). In general, however, individualization in the context of precision medicine needs at least an individual MRI and current flow calculations, e.g. using programs like SimNibs. Such “fingerprinting” of individuals, also on the basis of resting-state neuroimaging (e.g., MRI, fMRI, DTI) and electrophysiological (e.g., EEG) data is becoming a standard in neuroscience, with evidence of the possibility to accurately identify unique individual features of brain structure and function that are also of particular relevance for explaining interindividual variability in cognitive performance (Finn et al., 2015; Ozdemir et al., 2021). Individualized neuroenhancement solutions should start with selecting the target. For example, in a memory enhancement intervention that would mean selecting the most appropriate target region/network for the participant depending on his or her imaging/electrophysiology correlates of memory performance, and individualization of stimulation intensity based on parameters such as cortical excitability, biophysical modeling and other features obtained, e.g. by combined TMS-EEG recordings of TMS evoked potential amplitudes for non-motor regions. Recent retrospective studies have shown the importance of proper selection of individual rTMS targets within the left DLPFC on the basis of fMRI functional connectivity patterns in patients with major depression (Siddiqi et al., 2021). Evidence is also available for genetic factors that predict the response to stimulation, such as BDNF polymorphisms (Jannati et al., 2017), as well as for baseline cognitive performance (Krause & Cohen Kadosh, 2014). Efforts have also been made to prospectively tailor stimulation parameters on the

basis of individual brain or cognitive features, for instance by personalizing the stimulation frequency of tACS by targeting the individual alpha frequency (Kasten et al. , 2016), tailoring stimulation frequency and intensity of tACS based on cognitive ability (van Bueren et al. , 2021) or by identifying optimal TMS targets via analysis of individual connectome data. Overall, the better results with personalized neuroenhancement approaches seem to be widely accepted (Horn & Fox, 2020), and the required technology is partially ready including, e.g., portable EEG headsets that will enable remote brain activity recording and treatment adjustment, wearables to track physiological parameters (e.g., heart rate, sleep patterns), as well as analytical methods for the creation of individual target maps accounting for differences in anatomy and function (Kasten et al. , 2019; Ruffini et al., 2020).

However, questions remain. One aspect to consider is the temporal framework of personalization approaches and the issue of *state- vs trait-based* personalization. While genetic features constitute a trait since they are unlikely to change, changes in functional brain features, such as EEG activity and fMRI connectivity, are to be expected over the course of weeks, days, hours and even shorter intervals. Optimizing for individual features known to be fluctuating over time might not necessarily lead to an increase in effectiveness but actually introduce more variable effects between or within sessions, unless repeated baseline assessments are performed, or real-time time manipulation of brain stimulation parameters is achieved via simultaneous brain activity recording (e.g., closed-loop EEG-tES or EEG-TMS)(Zrenner et al., 2020). Related to this issue is the aspect of the generalizability of effects in larger populations. This is particularly relevant when promoting novel therapeutic solutions. Paradoxically, the adoption of a normative, population-level template describing, for instance, the average patterns of activity during a memory task recorded in a large sample of patients with mild-to-moderate AD, might constitute the optimal balance between personalization and generalizability, by increasing the chance that patients with similar conditions will benefit from the proposed stimulation approach even though it might be slightly sub-optimal at the individual level. Finally, barriers to the adoption of technology and the approval of novel clinical interventions should also be considered, including costs and burden of exposing patients to additional data acquisition sessions (e.g., fMRI, PET, TMS-EEG). In summary, neuroenhancement will benefit greatly from a better understanding of individual characteristics and from the development of the necessary technology to increase temporal and spatial resolution of individual “fingerprints”. A careful cost-benefit analysis is

also necessary to prevent unnecessary complexity that might prevent large-scale adoption of neuroenhancement interventions.

### **15. The media hype about neuroenhancement and brain stimulation**

The phenomenon of hyperbolic media reports on neurotechnology or neuroscience in general is not new (see e.g., Racine et al. , 2017). However, the proliferation of such reports and the explicit link of neurotechnology with prospects of enhancement, whether in the form of military funding for research that might provide an edge, or in terms of futuristic visions of a “posthuman” society, can generate a backlash. Namely, an overarchingly positive (“Hype and Hope”) attitude towards enhancement was met head on by an equally broad negative (“Doom and Gloom”) attitude (see Voarino et al. , 2016). However, public interests and democratic deliberation are best served if new neurostimulation technologies are viewed not through any one all-inclusive lens but on a case-by-case basis, so that self-regulation of the academic community and industry standards can be accomplished in benign cases, whereas for more serious cases, regulation could be approached by timely “agenda setting.” Polls conducted by social scientists have tracked what the general public considers to be “most important problems” for society, and research has shown that such “important problems” tend to closely follow media coverage (Birkland, 2005).

Early media discussions about neurostimulation have been optimistic, with little explicit discussion concerning ethical issues, therapeutic limitations, or adverse effects that could result from their wide-spread use (Dubljevic et al. , 2014, Racine et al. , 2007). Racine and colleagues analyzed print media coverage on neurostimulation between 1994 and 2004, and reported a distorted depiction, which often included first person narratives of patients, and sometimes celebrities, resembling “miracle stories.” The media used buzzwords or short catchy phrases, such as “currents of hope” and “magnetic appeal,” to convey the message that neurostimulation is a “scientific breakthrough.” The early reporting was fairly general, with only 20% (N=46) of the sample in Racine and colleagues’ study specifically referring to a NIBS technique such as TMS.

Further empirical work on media representation of brain stimulation was influenced by the appearance of commercial tDCS devices on the market, and by the specific use of enhancement claims in order to avoid FDA regulation of therapeutic devices. In 2014,

Dubljevic and colleagues reported strong and potentially misleading statements about the real-world effects and applicability of tDCS, even in otherwise serious news outlets (e.g., *“schoolchildren who struggle to grasp mathematics could benefit from having their brains roused with electricity”* [*The Guardian*, April 11, 2010]).

The concept of agenda-setting by the media (McCombs & Shaw 1972) describes how media coverage impacts what the general public considers to be a major issue. Bernard Cohen (Cohen, 1963) offered what is considered to be the classic summary of agenda setting: *“The press may not be successful much of the time in telling people what to think, but it is stunningly successful in telling its readers what to think about”* (p. 13). Issues that are most salient - defined by frequency, location, and/or length of coverage - in the media become part of the public discourse (Birkland, 2016). The more coverage an issue receives, the more the public pays attention (McCombs, Shaw, & Weaver, 2014). Media are further found to impact the public’s awareness of an issue based on *how* they report on a subject (McCombs & Shaw 1972).

While neurostimulation techniques could be therapeutically legitimate for a number of people with diverse health concerns, access to care may be influenced by misleading media portrayals that offer overly positive or overly negative perspectives. Positively biased information may lead to unrealistic expectations, which fail to materialize for the majority of people, ultimately leading individuals to see their results as a “let-down.” For example, media that paint a brain stimulation technique such as rTMS as “life-changing” or a “miracle cure” for depression may unwittingly set up those people who were treated and experienced only mild benefit, for disappointment (Dubljević et al., 2020). Similarly, as in situations of direct-to-consumer advertising, such stories could encourage patients to seek out treatment, in an era of self-diagnosis, that is not appropriate for them. On the other hand, negative perspectives may discourage patients from seeking or continuing treatment, even if the technology has legitimate, potential benefits. For example, overly negative testimonials or op-ed pieces from people who were disillusioned with NIBS might discourage others from seeking this treatment, even in cases where they could be suitable candidates. Given the implications of media representation for potential patients, media should take a balanced approach to discussing health technologies such as NIBS.

## **16. NIBS in military environments for performance enhancement**

Military organizations have been interested in NIBS for some time for its possible therapeutic use in treating diseases and injuries typical of military occupations (Oberman et al., 2020), but they have also recognized the usefulness, even the need, to explore cognitive enhancement for national security (Malish, 2017). It has been pointed out that with advancing technology, the modern soldier requires more complex mental skills than in the past, but that cognitive performance degrades in combat situations, and that NIBS may provide the means to counteract that degradation (Davis and Smith, 2019).

This interest was present early on: for instance, in 2002, the US Defense Advanced Research Projects Agency (DARPA) invested in using TMS for performance enhancement at a time when only a few TMS studies had even reported such effects, calling them “paradoxical,” as they were unexpected. As part of a larger effort to remediate the cognitive effects of sleep deprivation, a common problem in military situations, DARPA had funded a few groups to make the attempt using TMS. One group from Columbia University succeeded, by focusing specifically on working memory, a process central to executive cognitive function, in which deficits caused by sleep deprivation typically manifest as slowing of response and lapsing (missed trials) during the task. This was achieved in a succession of steps, beginning with a study investigating the experimental parameters that might lead to performance enhancement rather than disruption (Luber et al., 2007). Targeting nodes of the fronto-parietal executive network (not called that at the time) based on group fMRI activations of a delayed-match-to-sample (DMS) task, trains of rTMS to the parietal cortex during the delay period of the task at 5 Hz (but not 1 or 20 Hz) resulted in speeded performance compared with sham. This effect was reproduced in a different group of subjects again with stimulation during the delay phase of the task, but not during the test phase. In the next step, using an fMRI network for neuronavigated targeting that was activated by the DMS task but decreasing in its activation with sleep deprivation, an acute remediation of sleep deprivation-induced response slowing was seen in sleep deprived subjects for 5 Hz rTMS applied within the fMRI network, but not outside it (Luber et al., 2008). In a third step, this effect was prolonged well past the application of rTMS. Here, rTMS was applied in a number of sessions over the course of the two-day sleep deprivation period while the DMS task was performed. A day after the last rTMS was given, a group of sleep deprived subjects who had received sham stimulation showed the typical slowing and

lapsing deficits, while a group who had received active rTMS performed deficit-free (Luber et al., 2013). Overall, this was an early and concerted effort by a military research organization to develop NIBS applications useful in a military environment: a paradigm was developed, using fMRI-guidance, co-activation of targeted cortex using concurrent task performance and TMS, and multiple sessions to create a cumulative effect, which resulted in a long-lasting TMS-caused cognitive enhancement. While DARPA has gone on to study cognitive enhancement using a more invasive approach (e.g., DBS to affect memory: Deadwyler et al., 2017), other institutions such as the US National Institute on Aging adopted the NIBS paradigm to explore enhancement of working memory in aging populations (e.g., Beynel et al., 2019).

Military organizations have also investigated the use of tES for performance enhancement. For example, the U.S. Air Force Research Laboratory created a program to actively seek enhancement technologies to help bridge the gap between the ever-stronger computational capacities of machine interfaces and the limitations of their human operators (McKinley et al., 2012). For example, they sponsored a study that reported that tDCS applied while subjects were being trained to search complex simulated radar images of terrain containing buildings and vehicles for military targets increased their accuracy by about 25% over non-stimulation conditions, suggesting the potential usefulness of tDCS in accelerating training (McKinley et al., 2013). Another study they sponsored showed the efficacy of using tDCS to train personnel in threat detection using virtual reality software, where tDCS increased perceptual sensitivity in detecting threats across training sessions significantly compared with non-stimulation conditions (Parasuraman and Galster, 2013). A recent review by an Australian military research organization aggregated tDCS studies of enhancement by typical psychological categories such as attention, perception, memory, and reasoning, and also by communication skills, physical performance, and resilience, and discusses their potential benefits in the military (Davis and Smith, 2019). The review also addresses possible risks as well, and offers a number of ways to move forward, although another recent review cautions that at the current state of research, tES, while promising for military use, is far from ready for use in an operational environment (Feltman et al., 2020). Also important in this regard are ethical concerns involving the use of NIBS in the military—concerns about coercion and autonomy, and risks to others in battlefield conditions. These

concerns and related questions have also been discussed in previous reviews (Sehm and Ragert, 2013; Levasseur-Moreau et al., 2013).

## **17. Neuroenhancement and self-applied stimulation: more risks outside the lab environment? Neuroethics and neuroenhancement**

The non-professional or “lay” use of electrical stimulation has a long history dating back several centuries in both Europe and the United States (Kadosh, 2014). In the late nineteenth and early twentieth century, handheld devices known as “medical batteries” that provided either alternating or direct current were sold to both physicians and the public with claims of treating a wide variety of ailments and disorders (Peña 2003; Currier 2004; Waits 2013; Wexler 2017b). Aside from the marketing of a handful of products in the 1920s and 1930s for “rejuvenation” and “reinvigoration,” the notion of the use of electricity for enhancement—cognitive or otherwise—appears to have been largely absent from most historical marketing claims (Wexler 2017a).

The contemporary movement regarding the lay use of tES for enhancement began in 2011, when lay individuals began to construct tDCS devices in their homes. Since then, dozens of companies, largely based in the U.S., have marketed ready-to-wear tES devices for brain optimization and cognitive enhancement (Wexler 2015; McCall et al. 2019). To date, three empirical studies have been conducted to better understand users of self-directed home tES devices (Jwa 2015; Wexler 2016; Wexler 2018; for review see Wexler 2020). All found that most users were male and largely based in North America. Two earlier studies found that the typical user was in her/his 20s or 30s, whereas the more recent study (Wexler 2018) reported a mean age of 45. While some individuals used stimulation at home for treatment purposes (most commonly depression), they more frequently reported using tES for cognitive enhancement, specifically for improving focus and memory. Notably, however, those who have utilized tES for enhancement do not report it to be particularly successful (Wexler 2018).

Further research is needed on the potential of tES to produce neuroenhancement in otherwise healthy individuals (normally functioning brains) and outside of controlled laboratory environments, as well as on the potential for adverse events heretofore not identified in human trials. In addition, further research is warranted on if and when benefits

in one cognitive domain are at the expense of other cognitive functions, is not greater (Iuculano and Cohen Kadosh, 2013). Such considerations may be device or protocol specific. For example, one double-blind, sham-controlled study suggested that using a non-CE-marked commercial tDCS headset in healthy young adults produced detrimental effects on cognitive abilities instead of the promised ones, i.e., *“let the force of electricity excite your neurons into firing faster”* (Steenbergen et al., 2016).

The most likely adverse effect that may occur to inexperienced users is a skin irritation due to improper electrode impedance, missing electrolyte, tap water used as electrolyte, dried out sponge electrodes or application/removal of electrodes while the device is switched on (Antal et al., 2017; Woods et al., 2016). Even those lay users of tES devices, which do adequately provide for dose control (setting restrictions on their use), may fail to position the electrodes over the correct cortical target due to lack of anatomical knowledge, or may apply the wrong dosage with respect to tES intensity and duration, or may reverse stimulation polarity by switching anode and cathode, causing unintended effects (Fitz and Reiner, 2015) such as decrease performance in an intelligence test (Sellers et al., 2015). Furthermore, the time spent at home by an individual under stimulation may not be limited by some device (in contrast to remote-supervised use that provides for remote control; Charvet et al., 2020). This results in unknowns in regard to outcomes (Cohen Kadosh et al. 2013), in the attempt to “overclock” one’s own cognitive abilities, for example for improving online gaming performance (Santarnecci et al. 2013). Here, non-linearities of effects due to extended dosages, including stimulation intensity, duration, and inappropriate intervals in case of repeated stimulation might play a role (Agboada et al., 2020, Agboada et al., 2019, Batsikadze et al., 2013a, Monte-Silva et al., 2013, Mosayebi Samani et al., 2019, Mosayebi Samani et al., 2020). While there is no evidence tES or TMS can be addictive (and indeed some research of using tES / TMS to treat addictions) with broadening use cases, the potential for individuals (otherwise prone to addictive behavior) to excessive auto-stimulation should be studied (Santarnecci et al. 2013).

Many scholarly and media articles have portrayed the home use of tES as increasing, although an empirical assessment of the phenomenon in “real-world” scenarios is challenging. It is clear, however, that the home use of tES has not become mainstream at this time, but rather has remained a subculture. While devices continue to be sold—one

online survey of tES devices claims to have sold tens of thousands of products per year to consumers (Waltz, 2019)—the effectiveness of these devices, and the value they provide to consumers, remains an open question. Still, companies continue to bring new devices to market, in many regions aided by the lack of strict oversight from regulatory bodies regarding products marketed for enhancement or wellness purposes.

## 18. Neuroenhancement and regulating the marketing of devices

Most nations clearly differentiate the regulation of medical devices from that of other instruments and appliances. This distinction stems from the high standards for regulating the marketing of devices for medical diagnosis and treatment. While how medical devices are regulated is often determined by their perceived risk level, whether a device is considered to fall under medical device regulations (MDR) is not governed by risk, but rather it is designed to diagnose or treat a medical condition (e.g., a chainsaw would not be regulated as a medical device despite posing a clear potential health risk). Therefore, the perceived risk of a device cannot be the basis to decide if it should be regulated as a medical device. Rather, long-standing and well-developed regulations provide guidance on this topic.

Neuromodulation devices marketed for wellness and cognitive enhancement without explicit connection to a disease are not considered medical devices in all well-known U.S. jurisdictions. However, in some cases (notably the updated EU MDR) “*products without an intended medical purpose*” may fall under medical device regulations; in addition practices developed to ensure medical device quality (e.g., risk management) may be adopted through voluntary industry standards (Bikson et al., 2018). In the EU all tES & TMS devices have been treated as Medical Devices since May, 26<sup>th</sup> 2021; this includes any electrical, magnetic or electro-magnetic stimulation device for cognitive enhancements. The term *Medical Device Directive* (MDD) was taken out of action (beside the transition period of existing Medical Devices) by May, 25<sup>th</sup> 2021. According to the new MDR, tES and TMS devices will be regulated as stated in MDR ANNEX XVI (6) and at the official pages of the EU [https://ec.europa.eu/health/md\\_newregulations/getting\\_ready/manufacturers\\_devices\\_without\\_intended\\_medical\\_en](https://ec.europa.eu/health/md_newregulations/getting_ready/manufacturers_devices_without_intended_medical_en) & #3 [https://ec.europa.eu/health/sites/default/files/md\\_sector/docs/md\\_rolling-plan\\_en.pdf](https://ec.europa.eu/health/sites/default/files/md_sector/docs/md_rolling-plan_en.pdf).- However, the *Common Specifications* are not published yet. Those will give a clear

understanding how those devices should be verified & tested with respect to risks and safety.

When specific agencies are assigned to regulate medical devices (e.g., the FDA in the USA) they: 1) specifically regulate manufacturers and not the practice of medicine (e.g., physicians) or consumer behavior, and 2) focus on regulating the claims made by the manufacturer (also called the label) regarding specific medical uses. Medical devices are also regulated in how they can be sold, for example only to medical centers, to individuals under a prescription, or over-the-counter (OTC). There are currently hundreds of OTC products available, and in many cases the output of these OTC products exceed the output capacity of low intensity tES devices (Bikson et al., 2018). In the context of a regulated medical device, how instruments are accessed is critical and that a well-established framework exists for this purpose (e.g., human factors testing). For non-medical devices, any such restrictions are largely absent with few exceptions usually when there is an established and exceptionally high risk of harm (e.g., firearms, vehicles, illegal drugs). However, simply the potential for harm is not a basis for regulation (e.g., hammer and nails). Even in the few cases where special regulations exist for consumer devices, it is rarely under (subverting) MDRs. It would be an unheard-of situation, when physicians would be required to prescribe (or otherwise limit access to) a device that is not regulated as a medical device. The EU's updated MDR takes an explicit and special stance on "*Equipment intended for brain stimulation that apply electrical currents or magnetic or electromagnetic fields that penetrate the cranium to modify neuronal activity in the brain*", regulating such devices under MDR. Those devices are defined as Medical Devices without an Indented Use (Medical Claim).

As noted, governments (mainly outside the EU) only rarely restrict sales of non-medical devices to consumers, except for firearms and such. The theoretical potential for a non-medical device to be used by an individual for self-directed medical care has no bearing on how the device is regulated. For neuromodulation devices there may be a potential for both medical (e.g., insomnia) and wellness (e.g., "good night sleep") uses. This cannot in-itself justify regulation for the latter if there are no further, specific reasons. This raises the conundrum that a device, such as a tDCS device, that is marketed for wellness (thus not as a medical device) may be obtained by an individual and used for self-directed treatment.

It is therefore not correct to suggest that there is no legal guidance for consumer neuromodulation, at least in most of the countries. First, as noted above there is an extensive framework for distinguishing between medical devices and wellness devices. Indeed, in some cases the medical device regulator even provides an explicit mechanism to confirm that a device is not a medical device, and is thus not regulated as such (e.g., the 513g mechanism for the US FDA). Second, non-medical devices are subject to a range of regulations including those dealing with fair marketing (e.g., the Federal Communications Commission - FCC in the US). Third, in some regions additional regulations are applied to non-medical neuromodulation devices (e.g., in the EU).

While we think that companies that sell brain stimulation devices to the public should be based on scientific evidence to support their own product, rather than research done in the lab, with other devices, we acknowledge that any discussion on increasing the burden of regulation on companies or individual users must be conducted with an awareness of this spectrum of existing regulations and the rationale for both imposing and limiting regulatory burdens.

## **19. Conclusions and Recommendations**

While substantial research on low intensity tES effects on human learning and memory has already been conducted, recent reviews have noted substantial methodological variability, frequent lack of rigorous experimental control and overall highly variable outcomes within and between studies. Nonetheless, these studies have shown that there are potential positive effects on behavior and brain function, but the current “state-of-the-art” is not (yet) suited for translational application. This would require systematic and coordinated evaluation of a huge parameter space (type of stimulation, montage, dose of stimulation, number of repetition and breaks between them, and for the case of tACS, frequency - or the combination of frequencies - and phase), further clarification of the underlying neural mechanisms, and likely individual adaptation of interventions based on this knowledge. Per example, more (e.g., stimulation with higher intensity and for a longer time) might not be better (Gamboa et al. , 2010) and might even reverse the desired effects. Short-term as well as long-term rebound effects could be at play: homeostatic processes are an integral part of the brain’s function.

Furthermore, the translation of laboratory findings of selected aspects of cognitive or behavioral enhancement by NIBS to real-life contexts is largely unexplored. For instance, we

might observe improved performance in a working memory task after applying a form of NIBS in the laboratory, but there are no data if this would translate to improved performance at work or school when applied in settings outside the laboratory. Additional research on the effects of NIBS is needed, including how use and outcomes may vary in “real world” settings. Effects observed using restricted (and abstracted) laboratory tests may not generalize. Additional research is needed on the long-term effects of repeated NIBS sessions.

Reports in social media that promise dramatic improvements in cognitive functions may unduly influence healthcare providers, patients, and families when weighing the harm-benefit ratio and lead them to disregard risk factors that are being introduced by the application of neurostimulation. Besides this, tES inevitably induces placebo and nocebo effects in humans. This is closely related to the need to develop optimized sham protocols, e.g., based on electric field simulation (Neri et al., 2020) and appropriate evaluation methods (Fonteneau et al., 2019, Turner et al., 2020) to assure appropriate double blinding in forthcoming research and clinical studies.

To maximize reliability, tES devices marketed for cognitive enhancement should be engineered and manufactured to standards adopting appropriate practices established for medical devices, whether guided by industry standards (Bikson et al., 2018) or national regulators (as the new law in the EU). There will be Common Specifications as mentioned in EU’s rolling plan item #3 in [https://ec.europa.eu/health/sites/default/files/md\\_sector/docs/md\\_rolling-plan\\_en.pdf](https://ec.europa.eu/health/sites/default/files/md_sector/docs/md_rolling-plan_en.pdf) by next year. According to this, manufacturers have to adapt to these regulations within 6 months after publication. This will include tES devices for cognitive enhancement. Notwithstanding the tolerability and safety of tES as used in human trials (Antal et al., 2017; Bikson et al., 2016), efforts to provide tES directly to consumers (outside of remote-supervised use; Charvet et al., 2015) must consider the potential for improper use (Wurzman et al., 2016) which in turn can produce unpredictable outcomes.

The consensus among experts is that low-intensity tES is safe so long as accepted protocols (dose, inclusion/exclusion, proper electrode attachment) are followed. Devices that do not respect prior limitations on these factors, cannot claim to be “safe” based simply on these studies. Based on our previous recommendations (original document: [https://wfneurology.org/news\\_events/archived-news/2016-01-25-ifcn](https://wfneurology.org/news_events/archived-news/2016-01-25-ifcn)), we strongly

encourage that if low intensity tES is at all considered for neuroenhancement, the following conditions have to be taken into account:

- We recommend that tES in the treatment of a medical indication (at home or in the clinic) be done with a medical grade device and under on-site or remote supervision (Charvet et al., 2015) of a medical provider or trained personnel. The self-directed use of tES for medical indication - namely without a prescription or remote-supervision - requires device specific consideration of human factors and associated specific federal regulation. For example, in the United States, such devices are approved as Over-The-Counter (OTC).
- In medical applications, a necessary and sufficient condition is the use of protocols, which have been demonstrated in peer-reviewed clinical trials to be both safe and efficacious. The manufacturer should provide this information in the user manual, based on clinical evaluations. If a manufacturer makes any device claims that would be considered by relevant medical-device regulatory agencies as medical claims such that the device is a medical device, then evidently all standards for medical devices set by the regulatory agency will apply. In the case of non-medical claims (e.g., wellness or enhancement claims), regulation by medical-device agencies will vary. Manufacturers must follow ethical and best practices in marketing devices, including not misleading users by referencing effects from human trials that are not likely provided by the specific device.
- In non-medical environment (e.g., at home), safety should also be standard. The published effects of tES on circumscribed cognitive processes can vary, which in some cases is attributable to nuanced differences in protocols (e.g., dose, task difficulty, subject state and prior training). If a given 'home-applied' protocol has not been formally tested, and many sensitive parameters not carefully reproduced, stimulation results are unpredictable. If a self-directed device and a self-controlled protocol (the subject decides for what kind of application, using which montage and stimulation duration) do not meaningfully copy the factors tested in human trials (e.g., the inclusions/exclusion criteria, the tested population or dose of the stimulation) then the outcomes of such stimulation may not be the same as in those trials. This means that such devices cannot readily claim the benefits of those trials.

- Regardless of intended use, we recommend that self-directed non-invasive brain stimulation devices be designed, produced, and distributed following engineering risk-management procedures, for example as outlined in the LOTES-2017 industry guidance and/or mandated by regulatory agencies such as the revised EU CE guidance. We do not suggest that all consumer brain stimulation devices must be regulated fully as medical devices and indeed such a recommendation cannot be easily implemented since many government agencies tasked with regulating medical devices do not explicitly regulate "wellness-use" devices.
- We also recommend continuous refinements of education and certification programs for researchers and clinicians employing brain stimulation methods worldwide, including workshops and education material, etc. This should address the potential irresponsible and unsafe use of NIBS methods, and the establishment of requirements regarding who can use these methods and under which circumstances, including ethical governance.

## References

- Abdelmoula, A., Baudry, S., and Duchateau, J. (2016). Anodal transcranial direct current stimulation enhances time to task failure of a submaximal contraction of elbow flexors without changing corticospinal excitability. *Neuroscience* 322, 94–103.
- Aberra AS, Wang B, Grill WM, Peterchev AV. Simulation of transcranial magnetic stimulation in head model with morphologically-realistic cortical neurons. *Brain stimulation*. 2020;13:175-89.
- Agboada D, Mosayebi Samani M, Jamil A, Kuo MF, Nitsche MA. Expanding the parameter space of anodal transcranial direct current stimulation of the primary motor cortex. *Scientific reports*. 2019;9:18185.
- Agboada D, Mosayebi-Samani M, Kuo MF, Nitsche MA. Induction of long-term potentiation-like plasticity in the primary motor cortex with repeated anodal transcranial direct current stimulation - Better effects with intensified protocols? *Brain stimulation*. 2020;13:987-97.
- Aihara, T., Ogawa, T., Shimokawa, T., & Yamashita, O. (2017). Anodal transcranial direct current stimulation of the right anterior temporal lobe did not significantly affect verbal insight. *PloS One*, 12(9), e0184749. <https://doi.org/10.1371/journal.pone.0184749>
- Almquist, J. N.-F., Mathan, S., Brem, A.-K., Plessow, F., McKanna, J., Santarnecchi, E., Pascual-Leone, A., Cohen Kadosh, R., Pavel, M., & Yeung, N. (2019). FAST: A Novel, Executive Function-Based Approach to Cognitive Enhancement. *Frontiers in Human Neuroscience*, 13, 235. <https://doi.org/10.3389/fnhum.2019.00235>
- Ambrus GG, Al-Moyed H, Chaieb L, Sarp L, Antal A, Paulus W. The fade-in - Short stimulation - Fade out approach to sham tDCS - Reliable at 1 mA for naïve and experienced subjects, but not investigators. *Brain Stimul* 2012;5:499–504. <https://doi.org/10.1016/j.brs.2011.12.001>.
- Andrade AC, Magnavita GM, Allegro JV, Neto CE, Lucena Rde C, Fregni F. Feasibility of transcranial direct current stimulation use in children aged 5 to 12 years. *Journal of child neurology*. 2014;29:1360-5.
- Andrews, S. C., Hoy, K. E., Enticott, P. G., Daskalakis, Z. J., & Fitzgerald, P. B. (2011). Improving working memory: The effect of combining cognitive activity and anodal transcranial direct current stimulation to the left dorsolateral prefrontal cortex. *Brain Stimulation*, 4(2), 84–89. <https://doi.org/10.1016/j.brs.2010.06.004>
- Angius, L., Hopker, J. G., Marcora, S. M., and Mauger, A. R. (2015). The effect of transcranial direct current stimulation of the motor cortex on exercise-induced pain. *Eur. J. Appl. Physiol.* 115, 2311–2319.
- Angius, L., Mauger, A. R., Hopker, J., Pascual-Leone, A., Santarnecchi, E., and Marcora, S. M. (2018). Bilateral extracephalic transcranial direct current stimulation improves endurance performance in healthy individuals. *Brain Stimul.* 11, 108–117.
- Antal A, Alekseichuk I, Bikson M, Brockmüller J, Brunoni AR, Chen R, Cohen LG, Douthwaite G, Ellrich J, Flöel A, Fregni F, George MS, Hamilton R, Hauelsen J, Herrmann CS, Hummel FC, Lefaucheur JP, Liebetanz D, Loo CK, McCaig CD, Miniussi C, Miranda PC, Moliadze V, Nitsche MA, Nowak R, Padberg F, Pascual-Leone A, Poppendieck W, Priori A, Rossi S, Rossini PM, Rothwell J, Rueger MA, Ruffini G, Schellhorn K, Siebner HR, Ugawa Y, Wexler A, Ziemann U, Hallett M, Paulus W. Low intensity transcranial electric stimulation: Safety, ethical, legal regulatory and application guidelines. *Clin Neurophysiol*. 2017 Sep;128(9):1774-1809.
- Antal A, Boros K, Poreisz C, Chaieb L, Terney D, Paulus W. Comparatively weak after-effects of transcranial alternating current stimulation (tACS) on cortical excitability in humans. *Brain stimulation*. 2008;1:97-105.
- Antal A, Terney D, Poreisz C, Paulus W. Towards unravelling task-related modulations of neuroplastic changes induced in the human motor cortex. *The European journal of neuroscience*. 2007;26:2687-91.
- Antonenko D, Grittner U, Saturnino G, Nierhaus T, Thielscher A, Flöel A. Inter-individual and age-dependent variability in simulated electric fields induced by conventional transcranial electrical stimulation. *Neuroimage*. 2021;224:117413.
- Ardolino G, Bossi B, Barbieri S, Priori A. Non-synaptic mechanisms underlie the after-effects of cathodal transcutaneous direct current stimulation of the human brain. *The Journal of physiology*. 2005;568:653-63.
- Asamoah, B., Khatoun, A., & Mc Laughlin, M. (2019b) tACS motor system effects can be caused by transcutaneous stimulation of peripheral nerves. *Nature Communications*, 10, 266.
- Asseondi, S., Hu, R., Eskes, G., Pan, X., Zhou, J., & Shapiro, K. (2021). Impact of tDCS on working memory training is enhanced by strategy instructions in individuals with low working memory capacity. *Scientific Reports*, 11(1), 5531. <https://doi.org/10.1038/s41598-021-84298-3>

- Bandeira ID, Lins-Silva DH, Barouh JL, Faria-Guimaraes D, Dorea-Bandeira I, Souza LS, et al. Neuroplasticity and non-invasive brain stimulation in the developing brain. *Progress in brain research*. 2021;264:57-89.
- Banissy MJ, Muggleton NG. Transcranial direct current stimulation in sports training: potential approaches. *Front Hum Neurosci* 2013 (7):1-3.
- Bard I, Gaskell G, Allansdottir A, da Cunha RV, Eduard P, Hampel J, Hildt E, Hofmaier C, Kronberger N, Laursen S, Meijknecht A, Nordal S, Quintanilha A, Revuelta G, Saladié N, Sándor J, Santos JB, Seyringer S, Singh I, Somsen H, Toonders W, Torgersen H, Torre V, Varju M, Zwart H. Bottom Up Ethics - Neuroenhancement in Education and Employment. *Neuroethics*. 2018;11(3):309-322.
- Barron, F. (1955). The disposition toward originality. *Journal of Abnormal Psychology*, 51(3), 478–485. <https://doi.org/10.1037/h0048073>
- Barwood MJ, Butterworth J, Goodall S, House JR, Laws R, Nowicky A, et al. The effects of direct current stimulation on exercise performance, pacing and perception in temperate and hot environments. *Brain Stimul* 2016 (9):842-849.
- Batsikadze G, Moliadze V, Paulus W, Kuo MF, Nitsche MA. Partially non-linear stimulation intensity-dependent effects of direct current stimulation on motor cortex excitability in humans. *The Journal of physiology*. 2013;591:1987-2000.
- Benedetti, F., Carlino, E., & Pollo, A. (2011). How placebos change the patient's brain. *Neuropsychopharmacology*, 36(1), 339-354. doi:10.1038/npp.2010.81
- Berger, I., Dakwar-Kawar, O., Grossman, E. S., Nahum, M., & Cohen Kadosh, R. (2019). Scaffolding the Attention-Deficit/Hyperactivity Disorder Brain Using Random Noise Stimulation. *medRxiv*, 19005983. doi:10.1101/19005983
- Beynel, L., Davis, S.W. Crowell, C.A.C., Hilbig, S.A., Lim, W., Palmer, H., Brito, A., Peterchev, A.V., Luber, B., Lisanby, S.H., Cabeza, R., Appelbaum, L.G. (2019) Online repetitive transcranial magnetic stimulation during working memory in younger and older adults. *Plos One*;14(3):e0213707.
- Bikson M, Paneri B, Mourdoukoutas A, Esmailpour Z, Badran BW, Azzam R, Adair D, Datta A, Fang XH, Wingeier B, Chao D, Alonso-Alonso M, Lee K, Knotkova H, Woods AJ, Hagedorn D, Jeffery D, Giordano J, Tyler WJ. Limited output transcranial electrical stimulation (LOTES-2017): Engineering principles, regulatory statutes, and industry standards for wellness, over-the-counter, or prescription devices with low risk. *Brain Stimul*. 2018;11(1):134-157
- BinDawood A, Dickinson A, Aytumur A, Howarth C, Milne E, Jones M. Investigating the effects of tDCS on Visual Orientation Discrimination Task Performance: “The possible influence of placebo”. *J Cogn Enhanc* 2020;4:235–49. <https://doi.org/10.1007/s41465-019-00154-3>.
- Bindman LJ, Lippold OC, Redfearn JW. The Action of Brief Polarizing Currents on the Cerebral Cortex of the Rat (1) during Current Flow and (2) in the Production of Long-Lasting after-Effects. *The Journal of physiology*. 1964;172:369-82.
- Birkland, T. (2016). *An introduction to the policy process*, 4<sup>th</sup> edition, New York, NY: Routledge
- Birkland, T. A. (2005). *An introduction to the policy process : theories, concepts, and models of public policy making* (2nd ed.). Armonk, N.Y.: M.E. Sharpe.
- Boggio PS, Zaghi S, Fregni F. Modulation of emotions associated with images of human pain using anodal transcranial direct current stimulation (tDCS). *Neuropsychologia*. 2009;47:212-7.
- Boroda, E., Krueger, A. M., Bansal, P., Schumacher, M. J., Roy, A. V., Boys, C. J., Lim, K. O., & Wozniak, J. R. (2020). A randomized controlled trial of transcranial direct-current stimulation and cognitive training in children with fetal alcohol spectrum disorder. *Brain Stimulation*, 13(4), 1059–1068. <https://doi.org/10.1016/j.brs.2020.04.015>
- Bortoletto M, Pellicciari MC, Rodella C, Miniussi C. The interaction with task-induced activity is more important than polarization: a tDCS study. *Brain stimulation*. 2015;8:269-76.
- Brauer H, Kadish NE, Pedersen A, Siniatchkin M, Moliadze V. No Modulatory Effects when Stimulating the Right Inferior Frontal Gyrus with Continuous 6 Hz tACS and tRNS on Response Inhibition: A Behavioral Study. *Neural plasticity*. 2018;2018:3156796.

- Brem, A. K., Fried, P. J., Horvath, J. C., Robertson, E. M., & Pascual-Leone, A. (2014). Is neuroenhancement by noninvasive brain stimulation a net zero-sum proposition? *Neuroimage*, *85*, 1058-1068.
- Brem, A.-K., Almquist, J. N.-F., Mansfield, K., Plessow, F., Sella, F., Santarnecchi, E., Orhan, U., McKanna, J., Pavel, M., Mathan, S., Yeung, N., Pascual-Leone, A., Kadosh, R. C., & Honeywell SHARP Team authors. (2018). Modulating fluid intelligence performance through combined cognitive training and brain stimulation. *Neuropsychologia*, *118*(Pt A), 107–114.
- Brignani D, Ruzzoli M, Mauri P, Miniussi C. Is transcranial alternating current stimulation effective in modulating brain oscillations? *PLoS One*. 2013;8(2):e56589.
- Brunoni AR, Vanderhasselt MA, Boggio PS, Fregni F, Dantas EM, Mill JG, et al. Polarity- and valence-dependent effects of prefrontal transcranial direct current stimulation on heart rate variability and salivary cortisol. *Psychoneuroendocrinology*. 2013;38:58-66.
- Brunoni AR, Vanderhasselt MA. Working memory improvement with non-invasive brain stimulation of the dorsolateral prefrontal cortex: a systematic review and meta-analysis. *Brain Cogn*. 2014 Apr;86:1-9.
- Buch ER, Santarnecchi E, Antal A, Born J, Celnik PA, Classen J, et al. Effects of tDCS on motor learning and memory formation: A consensus and critical position paper. *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology*. 2017;128:589-603.
- Bueno-Lopez A, Eggert T, Dorn H, Danker-Hopfe H. Slow oscillatory transcranial direct current stimulation (so-tDCS) during slow wave sleep has no effects on declarative memory in healthy young subjects. *Brain stimulation*. 2019;12:948-58.
- Caligiore D, Pessulo G, Baldassarre G, Bostan AC, Strick PL, DOya K, Helmich RC, Dirx M, Houk J, Jorntell H, Lago-Rodeiguez A, Galea JM, Miall RC, Popa T, Kishore A, Verschure PFMJ, Zucca R, Herreros I. Consensus Paper: Towards a Systems-Level View of Cerebellar Function: the Interplay Between Cerebellum, Basal Ganglia, and Cortex. *Cerebellum* (2017) *16*:203–229.
- Cappelletti M, Gessaroli E, Hithersay R, Mitolo M, Didino D, Kanai R, et al. Transfer of cognitive training across magnitude dimensions achieved with concurrent brain stimulation of the parietal lobe. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2013;33:14899-907.
- Cappelletti, S., Piacentino, D., Sani, G., & Aromatario, M. (2015). Caffeine: cognitive and physical performance enhancer or psychoactive drug? *Curr Neuropharmacol*, *13*(1), 71-88. doi:10.2174/1570159X13666141210215655
- Castelo-Branco L, Fregni F. Home-Based Transcranial Direct Current Stimulation (tDCS) to Prevent and Treat Symptoms Related to Stress: A Potential Tool to Remediate the Behavioral Consequences of the COVID-19 Isolation Measures? *Frontiers in integrative neuroscience*. 2020;14:46.
- Cecere R, Rees G, Romei V (2015) Individual differences in alpha frequency drive crossmodal illusory perception. *Current Biology*, *25*, 231–235.
- Cellini N, Mednick SC. Stimulating the sleeping brain: Current approaches to modulating memory-related sleep physiology. *Journal of neuroscience methods*. 2019;316:125-36.
- Cellini N, Shimizu RE, Connolly PM, Armstrong DM, Hernandez LT, Polakiewicz AG, et al. Short Duration Repetitive Transcranial Electrical Stimulation During Sleep Enhances Declarative Memory of Facts. *Frontiers in human neuroscience*. 2019;13:123.
- Cerruti, C., & Schlaug, G. (2009). Anodal transcranial direct current stimulation of the prefrontal cortex enhances complex verbal associative thought. *Journal of Cognitive Neuroscience*, *21*(10), 1980–1987. <https://doi.org/10.1162/jocn.2008.21143>
- Charvet LE, Kasschau M, Datta A, Knotkova H, Stevens MC, Alonzo A, Loo C, Krull KR, Bikson M. *Front Syst Neurosci*. Remotely-supervised transcranial direct current stimulation (tDCS) for clinical trials: guidelines for technology and protocols. 2015;9:26
- Charvet LE, Shaw MT, Bikson M, Woods AJ, Knotkova H. *Brain Stimul*. Supervised transcranial direct current stimulation (tDCS) at home: A guide for clinical research and practice. 2020, *13*(3):686-693
- Chrysikou, E. G., Morrow, H. M., Flohrschutz, A., & Denney, L. (2021). Augmenting ideational fluency in a creativity task across multiple transcranial direct current stimulation montages. *Scientific Reports*, *11*(1), 8874. <https://doi.org/10.1038/s41598-021-85804-3>

- Chu CS, Li CT, Brunoni AR, Yang FC, Tseng PT, Tu YK, Stubbs B, Carvalho AF, Thompson T, Rajji TK, Yeh TC, Tsai CK, Chen TY, Li DJ, Hsu CW, Wu YC, Yu CL, Liang CS. Cognitive effects and acceptability of non-invasive brain stimulation on Alzheimer's disease and mild cognitive impairment: a component network meta-analysis. *J Neurol Neurosurg Psychiatry*. 2021 Feb;92(2):195-203.
- Clayton MS, Yeung N, Cohen Kadosh R (2019) Electrical stimulation of alpha oscillations stabilizes performance on visual attention tasks. *J Exp Psychol Gen* 148(2):203-220.
- Coffman, B. A., Clark, V. P., and Parasuraman, R. (2014) Battery powered thought: enhancement of attention, learning, and memory in healthy adults using transcranial direct current stimulation. *NeuroImage* 85, 895–908.
- McCombs, M. E., & Shaw, D. L. (1972). The agenda-setting function of mass media. *Public Opinion Quarterly*, 36, 176-187.
- Cogiamanian, F., Marceglia, S., Ardolino, G., Barbieri, S., and Priori, A. (2007). Improved isometric force endurance after transcranial direct current stimulation over the human motor cortical areas. *Eur. J. Neurosci*. 26, 242–249.
- Cohen, B. C. (1963). *The press and foreign policy*. Princeton, N.J.: Princeton University Press.
- Colagiuri B, Livesey EJ, Harris JA. Can expectancies produce placebo effects for implicit learning? *Psychon Bull Rev* 2011;18:399–405. <https://doi.org/10.3758/s13423-010-0041-1>.
- Colzato LS, Hommel B, Beste C. The Downsides of Cognitive Enhancement. *Neuroscientist*. 2020 Jul 30:1073858420945971.
- Colzato LS, Nitsche MA, Kibele A. Noninvasive brain stimulation and neural entrainment enhance athletic performance: a review. *J Cogn Enhanc* 2017 (1):73-79.
- Corp, D. T., Bereznicki, H. G. K., Clark, G. M., Youssef, G. J., Fried, P. J., Jannati, A., Davies, C. B., Gomes-Osman, J., Stamm, J., Chung, S. W., Bowe, S. J., Rogasch, N. C., Fitzgerald, P. B., Koch, G., Di Lazzaro, V., Pascual-Leone, A., & Enticott, P. G. (2020). Large-scale analysis of interindividual variability in theta-burst stimulation data: Results from the 'Big TMS Data Collaboration.' *Brain Stimulation*, 13(5), 1476–1488.
- Creutzfeldt OD, Fromm GH, Kapp H. Influence of transcortical d-c currents on cortical neuronal activity. *Experimental neurology*. 1962;5:436-52.
- Currier, D. P. 2004. *Guide to Electrotherapy Instruments and History of Their American Makers*. Infinity Publishing. Infinity Publishing.
- Daubner J, Arshaad MI, Henseler C, Hescheler J, Ehninger D, Broich K, et al. Pharmacological Neuroenhancement: Current Aspects of Categorization, Epidemiology, Pharmacology, Drug Development, Ethics, and Future Perspectives. *Neural plasticity*. 2021;2021:8823383.
- Davis NJ, Gold E, Pascual-Leone A, Bracewell RM. Challenges of proper placebo control for non-invasive brain stimulation in clinical and experimental applications. *Eur J Neurosci* 2013;38:2973–7. <https://doi.org/10.1111/ejn.12307>.
- Davis NJ. Neurodoping: brain stimulation as a performance enhancing measure. *Sports Med* 2013 (43):649-653.
- Davis NJ. Transcranial stimulation of the developing brain: a plea for extreme caution. *Frontiers in human neuroscience*. 2014;8:600.
- Davis, S.E. and Smith, G.A. (2029) Transcranial direct current stimulation use in warfighting: Benefits, risks, and future prospects. *Frontiers in Human Neuroscience*, 13: Article 114. doi: 10.3389/fnhum.2019.00114
- de Sousa AVC, Grittner U, Rujescu D, Külzow N, Flöel A. de Sousa AVC, et al Impact of 3-Day Combined Anodal Transcranial Direct Current Stimulation-Visuospatial Training on Object-Location Memory in Healthy Older Adults and Patients with Mild Cognitive Impairment. *J Alzheimers Dis*. 2020;75(1):223-244.
- Deadwyler, S.A., Hampson, R.E., Song, D., Opris, I., Gerhardt, G.A., Marmarelis, V.Z., Berger, T.W. (2017) A cognitive prosthesis for memory facilitation by closed-loop functional ensemble stimulation of hippocampal neurons in primate brain. *Experimental Neurology*, 287: 452-460.
- Doruk Camsari D, Kirkovski M, Croarkin PE. Therapeutic Applications of Noninvasive Neuromodulation in Children and Adolescents. *Psychiatr Clin North Am*. 2018;41:465-77.
- Dubljevic V, Saigle V, Racine E. The rising tide of tDCS in the media and academic literature. *Neuron*. 2014;82:731-6.

Dubljević, V., Coates McCall, I., & Illes, J. (2020). Neuroenhancement at work: Addressing the ethical, legal and social implications. In J. Trempe-Martineau & E. Racine (Eds.), *Organizational Neuroethics*. Cham, Switzerland: Springer. In Press.

Dubreuil-Vall, L., Gomez-Bernal, F., Villegas, A. C., Cirillo, P., Surman, C., Ruffini, G., Widge, A. S., & Camprodon, J. A. (2021). Transcranial Direct Current Stimulation to the Left Dorsolateral Prefrontal Cortex Improves Cognitive Control in Patients With Attention-Deficit/Hyperactivity Disorder: A Randomized Behavioral and Neurophysiological Study. *Biological Psychiatry: Cognitive Neuroscience and Neuroimaging*, *6(4)*, 439–448.

Eggert T, Dorn H, Sauter C, Nitsche MA, Bajbouj M, Danker-Hopfe H. No effects of slow oscillatory transcranial direct current stimulation (tDCS) on sleep-dependent memory consolidation in healthy elderly subjects. *Brain stimulation*. 2013;6:938-45.

Ehrhardt, S. E., Filmer, H. L., Wards, Y., Mattingley, J. B., & Dux, P. E. (2021). The influence of tDCS intensity on decision-making training and transfer outcomes. *Journal of Neurophysiology*, *125(2)*, 385–397. <https://doi.org/10.1152/jn.00423.2020>

Eilam-Stock T, George A, Lustberg M, Wolintz R, Krupp LB, Charvet LE. Telehealth transcranial direct current stimulation for recovery from Post-Acute Sequelae of SARS-CoV-2 (PASC). *Brain stimulation*. 2021;14:1520-2.

Ekhtiari H, Tavakoli H, Addolorato G, Baeken C, Bonci A, Campanella S, Castelo-Branco L, Challet-Bouju G, Clark VP, Claus E, Dannon PN, Del Felice A, den Uyl T, Diana M, di Giannantonio M, Fedota JR, Fitzgerald P, Gallimberti L, Grall-Bronnec M, Herremans SC, Herrmann MJ, Jamil A, Khedr E, Kouimtsidis C, Kozak K, Krupitsky E, Lamm C, Lechner WV, Madeo G, Malmir N, Martinotti G, McDonald WM, Montemitro C, Nakamura-Palacios EM, Nasehi M, Noël X, Nosratabadi M, Paulus M, Pettoruso M, Pradhan B, Prahara SK, Rafferty H, Sahlem G, Salmeron BJ, Sauvaget A, Schluter RS, Sergiou C, Shahbabaie A, Sheffer C, Spagnolo PA, Steele VR, Yuan TF, van Dongen JDM, Van Waes V, Venkatasubramanian G, Verdejo-García A, Verveer I, Welsh JW, Wesley MJ, Witkiewitz K, Yavari F, Zarrindast MR, Zawertailo L, Zhang X, Cha YH, George TP, Frohlich F, Goudriaan AE, Fecteau S, Daughters SB, Stein EA, Fregni F, Nitsche MA, Zangen A, Bikson M, Hanlon CA. Transcranial electrical and magnetic stimulation (tES and TMS) for addiction medicine: A consensus paper on the present state of the science and the road ahead. *Neurosci Biobehav Rev*. 2019 Sep;104:118-140. doi: 10.1016/j.neubiorev.2019.06.007.

Evans, C., Banissy, M. J., & Charlton, R. A. (2018). The efficacy of transcranial random noise stimulation (tRNS) on mood may depend on individual differences including age and trait mood. *Clin Neurophysiol*, *129(6)*, 1201-1208. doi:10.1016/j.clinph.2018.03.012

Fagerlund AJ, Freili JL, Danielsen TL, Aslaksen PM. No Effect of 2 mA Anodal tDCS Over the M1 on Performance and Practice Effect on Grooved Pegboard Test and Trail Making Test B. *eNeuro*. 2015;2.

Fekete T, Nikolaev AR, Knijf FD, Zharikova A, van Leeuwen C (2018) Multi-electrode alpha tACS during varying background tasks fails to modulate subsequent alpha power. *Frontiers in Neuroscience*, *12*.

Feltman, K.A., Hayes, A.M., Bernhardt, K.A., Nwala, E., and Kelly, A.M. (2020) Viability of tDCS in military environments for performance enhancement: A systematic review. *Military Medicine*, *185*: ½: e53-e60.

Ferreri F, Vecchio F, Ponzo D, Pasqualetti P, Rossini PM. Time-varying coupling of EEG oscillations predicts excitability fluctuations in the primary motor cortex as reflected by motor evoked potentials amplitude: an EEG-TMS study. *Human brain mapping*. 2014;35:1969-80.

Fertonani A, Brambilla M, Cotelli M, Miniussi C. The timing of cognitive plasticity in physiological aging: a tDCS study of naming. *Frontiers in aging neuroscience*. 2014;6:131.

Fertonani A, Ferrari C, Miniussi C. What do you feel if I apply transcranial electric stimulation? Safety, sensations and secondary induced effects. *Clin Neurophysiol* 2015;126:2181–8. <https://doi.org/10.1016/j.clinph.2015.03.015>.

Fertonani, A., Pirulli, C., & Miniussi, C. (2011). Random Noise Stimulation Improves Neuroplasticity in Perceptual Learning. *The Journal of Neuroscience*, *31(43)*, 15416-15423.

Fiani B, Zhu L, Musch BL, Briceno S, Andel R, Sadeq N, et al. The Neurophysiology of Caffeine as a Central Nervous System Stimulant and the Resultant Effects on Cognitive Function. *Cureus*. 2021;13:e15032.

Filmer, H. L., Lyons, M., Mattingley, J. B., & Dux, P. E. (2017). Anodal tDCS applied during multitasking training leads to transferable performance gains. *Scientific Reports*, *7(1)*, 12988. <https://doi.org/10.1038/s41598-017-13075-y>

- Finisguerra A, Borgatti R, Urgesi C. Non-invasive Brain Stimulation for the Rehabilitation of Children and Adolescents With Neurodevelopmental Disorders: A Systematic Review. *Frontiers in Psychology*. 2019;10.
- Finn, E. S., Shen, X., Scheinost, D., Rosenberg, M. D., Huang, J., Chun, M. M., Papademetris, X., & Constable, R. T. (2015). Functional connectome fingerprinting: Identifying individuals using patterns of brain connectivity. *Nat. Neurosci*, *18*(11), 1664–1671
- Fitz NS, Reiner PB. The challenge of crafting policy for do-it-yourself brain stimulation. *J Med Ethics* 2015, *41*(5): 410-412.
- Fonteneau C, Mondino M, Arns M, Baeken C, Bikson M, Brunoni AR, Burke MJ, Neuvonen T, Padberg F, Pascual-Leone A, Poulet E, Ruffini G, Santarnecchi E, Sauvaget A, Schellhorn K, Suaud-Chagny MF, Palm U, Brunelin J. Sham tDCS: A hidden source of variability? Reflections for further blinded, controlled trials. *Brain Stimul*. 2019;*12*(3):668-673
- Frank, B., Harty, S., Kluge, A., & Cohen Kadosh, R. (2018). Learning while multitasking: short and long-term benefits of brain stimulation. *Ergonomics*, *61*(11), 1454-1463.
- Fregni F, El-Hagrassy MM, Pacheco-Barrios K, Carvalho S, Leite J, Simis M, Brunelin J, Nakamura-Palacios EM, Marangolo P, Venkatasubramanian G, San-Juan D, Caumo W, Bikson M, Brunoni AR; Neuromodulation Center Working Group. Evidence-Based Guidelines and Secondary Meta-Analysis for the Use of Transcranial Direct Current Stimulation in Neurological and Psychiatric Disorders. *Int J Neuropsychopharmacol*. 2021;*24*(4):256-313.
- Frohlich F, Lustenberger C. Neuromodulation of sleep rhythms in schizophrenia: Towards the rational design of non-invasive brain stimulation. *Schizophrenia research*. 2020;*221*:71-80.
- Frohlich F, Sellers KK, Cordle AL. Targeting the neurophysiology of cognitive systems with transcranial alternating current stimulation. *Expert review of neurotherapeutics*. 2015;*15*:145-67.
- Furuya S, Klaus M, Nitsche MA, Paulus W, Altenmueller E. Ceiling Effects Prevent Further Improvement of Transcranial Stimulation in Skilled Musicians. *Journal of Neuroscience*. 2014;*34*:13834-9.
- Gamboa OL, Antal A, Moliadze V, Paulus W. Simply longer is not better: reversal of theta burst after-effect with prolonged stimulation. *Experimental brain research Experimentelle Hirnforschung Experimentation cerebrale*. 2010;*204*:181-7.
- Gandiga PC, Hummel FC, Cohen LG. Transcranial DC stimulation (tDCS): a tool for double-blind sham-controlled clinical studies in brain stimulation. *Clin Neurophysiol* 2006;*117*:845–50.  
<https://doi.org/10.1016/j.clinph.2005.12.003>.
- Gaskell G, Allansdottir A, Allum N, Castro P, Esmer Y, Fischler C, Jackson J, Kronberger N, Hampel J, Mejlgaard N, Quintanilha A, Rammer A, Revuelta G, Stares S, Torgersen H, Wager W. The 2010 Eurobarometer on the life sciences. *Nat Biotechnol*. 2011, *29*(2):113-4.
- Gill, J., Shah-Basak, P. P., & Hamilton, R. (2015). It's the Thought That Counts: Examining the Task-dependent Effects of Transcranial Direct Current Stimulation on Executive Function. *Brain Stimulation*, *8*(2), 253–259.  
<https://doi.org/10.1016/j.brs.2014.10.018>
- Gniss, S., Kappesser, J., & Hermann, C. (2020). Placebo effect in children: the role of expectation and learning. *Pain*, *161*(6), 1191-1201.
- Goder R, Baier PC, Beith B, Baecker C, Seeck-Hirschner M, Junghanns K, et al. Effects of transcranial direct current stimulation during sleep on memory performance in patients with schizophrenia. *Schizophrenia research*. 2013;*144*:153-4.
- Goering S, Klein E, Specker Sullivan L, Wexler A, Agüera Y Arcas B, Bi G, Carmena JM, Fins JJ, Friesen P, Gallant J, Huggins JE, Kellmeyer P, Marblestone A, Mitchell C, Parens E, Pham M, Rubel A, Sadato N, Teicher M, Wasserman D, Whittaker M, Wolpaw J, Yuste R. Recommendations for Responsible Development and Application of Neurotechnologies. *Neuroethics*. 2021 Apr *29*:1-22.
- Goldsworthy MR, Pitcher JB, Ridding MC. Neurorehabil Neural Repair. Spaced Noninvasive Brain Stimulation: Prospects for Inducing Long-Lasting Human Cortical Plasticity. 2015;*29*(8):714-21
- Grabner, R. H., Krenn, J., Fink, A., Arendasy, M., & Benedek, M. (2018). Effects of alpha and gamma transcranial alternating current stimulation (tACS) on verbal creativity and intelligence test performance. *Neuropsychologia*, *118*(Pt A), 91–98.
- Grandner MA. Sleep, Health, and Society. *Sleep Med Clin*. 2017;*12*:1-22.

- Green, A. E., Spiegel, K. A., Giangrande, E. J., Weinberger, A. B., Gallagher, N. M., & Turkeltaub, P. E. (2017a). Thinking Cap Plus Thinking Zap: TDCS of Frontopolar Cortex Improves Creative Analogical Reasoning and Facilitates Conscious Augmentation of State Creativity in Verb Generation. *Cerebral Cortex (New York, N.Y.: 1991)*, 27(4), 2628–2639.
- Greinacher R, Buhôt L, Möller L, Learmonth G. The time course of ineffective sham-blinding during low-intensity (1 mA) transcranial direct current stimulation. *Eur J Neurosci* 2019;50:3380–8. <https://doi.org/10.1111/ejn.14497>.
- Grimaldi D, Papalambros NA, Zee PC, Malkani RG. Neurostimulation techniques to enhance sleep and improve cognition in aging. *Neurobiology of disease*. 2020;141:104865.
- Grosprêtre S, Grandperrin Y, Nicolier M, Gimenez P, Vidal C, Tio G, Haffen E, Bennabi D Effect of transcranial direct current stimulation on the psychomotor, cognitive, and motor performances of power athletes. *Sci Rep*. 2021 May 6;11(1):9731.
- Grundey J, Amu R, Ambrus GG, Batsikadze G, Paulus W, Nitsche MA. Double dissociation of working memory and attentional processes in smokers and non-smokers with and without nicotine. *Psychopharmacology*. 2015;232:2491-501.
- Grundey J, Amu R, Batsikadze G, Paulus W, Nitsche MA. Diverging effects of nicotine on motor learning performance: Improvement in deprived smokers and attenuation in non-smokers. *Addict Behav*. 2017;74:90-7.
- Grundey J, Freznosa S, Klinker F, Lang N, Paulus W, Nitsche MA. Cortical excitability in smoking and not smoking individuals with and without nicotine. *Psychopharmacology*. 2013;229:653-64.
- Grundey J, Thirugnanasambandam N, Kaminsky K, Drees A, Skwirba AC, Lang N, et al. Neuroplasticity in cigarette smokers is altered under withdrawal and partially restituted by nicotine exposition. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2012a;32:4156-62.
- Grundey J, Thirugnanasambandam N, Kaminsky K, Drees A, Skwirba AC, Lang N, et al. Rapid effect of nicotine intake on neuroplasticity in non-smoking humans. *Frontiers in pharmacology*. 2012b;3:186.
- Guest, N. S., VanDusseldorp, T. A., Nelson, M. T., Grgic, J., Schoenfeld, B. J., Jenkins, N. D. M., . . . Campbell, B. I. (2021). International society of sports nutrition position stand: caffeine and exercise performance. *J Int Soc Sports Nutr*, 18(1), 1. doi:10.1186/s12970-020-00383-4
- Habich A, Fehér KD, Antonenko D, Boraxbekk CJ, Flöel A, Nissen C, Siebner HR, Thielscher A, Klöppel S. Stimulating aged brains with transcranial direct current stimulation: Opportunities and challenges. *Psychiatry Res Neuroimaging*. 2020 Dec 30;306:111179.
- Hamada M, Hanajima R, Terao Y, et al (2009) Primary motor cortical metaplasticity induced by priming over the supplementary motor area. *J Physiol* 587:4845–4862.
- Hamada M, Terao Y, Hanajima R, et al (2008) Bidirectional long-term motor cortical plasticity and metaplasticity induced by quadripulse transcranial magnetic stimulation. *J Physiol* 586:3927–3947.
- Hanajima R, Tanaka N, Tsutsumi R, Shirota Y, Shimizu T, Terao Y, et al. Effect of caffeine on long-term potentiation-like effects induced by quadripulse transcranial magnetic stimulation. *Experimental brain research Experimentelle Hirnforschung Experimentation cerebrale*. 2019;237:647-51.
- Harty, S., & Cohen Kadosh, R. (2019). Suboptimal Engagement of High-Level Cortical Regions Predicts Random-Noise-Related Gains in Sustained Attention. *Psychological Science*, 30(9), 1318-1332. doi:10.1177/0956797619856658
- Harty, S., Robertson, I. H., Miniussi, C., Sheehy, O. C., Devine, C. A., McCreery, S., & O'Connell, R. G. (2014). Transcranial Direct Current Stimulation over Right Dorsolateral Prefrontal Cortex Enhances Error Awareness in Older Age. *Journal of Neuroscience*, 34(10), 3646–3652. <https://doi.org/10.1523/JNEUROSCI.5308-13.2014>
- Hazime, F. A., da Cunha, R. A., Soliaman, R. R., Romancini, A. C. B., Pochini, A. D. C., Ejnisman, B., et al. (2017). Anodal transcranial direct current stimulation (TDCS) increases isometric strength of shoulder rotators muscles in handball players. *Int. J. Sports Phys. Ther.* 12, 402–407.
- Hean W, Fong Po-Yu, Leung TW, Huan Y. Protocols of non-invasive brain stimulation for neuroplasticity induction. *Neurosci Letters* 719 (2020) 1334-1337
- Helfrich RF, Knepper H, Nolte G, Struber D, Rach S, Herrmann CS, et al. Selective modulation of interhemispheric functional connectivity by HD-tACS shapes perception. *PLoS biology*. 2014a;12:e1002031.

- Helfrich RF, Schneider TR, Rach S, Trautmann-Lengsfeld SA, Engel AK, Herrmann CS (2014)b. Entrainment of Brain Oscillations by Transcranial Alternating Current Stimulation. *Current Biology* 24(3), 333-339.
- Henin S, Borges H, Shankar A, Sarac C, Melloni L, Friedman D, et al. Closed-Loop Acoustic Stimulation Enhances Sleep Oscillations But Not Memory Performance. *eNeuro*. 2019;6.
- Herrmann CS, Rach S, Neuling T, Strüber D (2013) Transcranial alternating current stimulation: a review of the underlying mechanisms and modulation of cognitive processes. *Frontiers in Human Neuroscience* 7:279.
- Hertenstein, E., Waibel, E., Frase, L., Riemann, D., Feige, B., Nitsche, M. A., Kaller, C. P., & Nissen, C. (2019). Modulation of creativity by transcranial direct current stimulation. *Brain Stimulation*, 12(5), 1213–1221. <https://doi.org/10.1016/j.brs.2019.06.004>
- Hikosaka O, Rand MK, Miyachi S, Miyashita K. Learning of sequential movements in the monkey: process of learning and retention of memory. *J Neurophysiol* 1995; 74: 1652-1661.
- Hikosaka O, Sakai K, Miyauchi S, Takino R, Sasaki Y, Pütz B. Activation of human presupplementary motor area in learning of sequential procedures: a functional MRI study. *J Neurophysiol* 1996; 76: 617-621.
- Hilgetag, C.C., Théoret, H., Pascual-Leone, A., 2001. Enhanced visual spatial attention ipsilateral to rTMS-induced “virtual lesions” of human parietal cortex. *Nat. Neurosci.* 4, 953–957.
- Horn, A., & Fox, M. D. (2020). Opportunities of connectomic neuromodulation. *NeuroImage*, 221, 117180.
- Horne, K. S., Filmer, H. L., Nott, Z. E., Hawi, Z., Pugsley, K., Mattingley, J. B., & Dux, P. E. (2021). Evidence against benefits from cognitive training and transcranial direct current stimulation in healthy older adults. *Nature Human Behaviour*, 5(1), 146–158. <https://doi.org/10.1038/s41562-020-00979-5>
- Huang YZ, Sommer M, Thickbroom G, Hamada M, Pascual-Leone A, Paulus W, et al. Consensus: New methodologies for brain stimulation. *Brain stimulation*. 2009;2:2-13.
- Hunold A, Haueisen J, Freitag CM, Siniatchkin M, Moliadze V. Cortical current density magnitudes during transcranial direct current stimulation correlate with skull thickness in children, adolescent and young adults. *Progress in brain research*. 2021;264:41-56.
- Huo L, Zhu X, Zheng Z, Ma J, Ma Z, Gui W, Li J. Effects of Transcranial Direct Current Stimulation on Episodic Memory in Older Adults: A Meta-analysis. *J Gerontol B Psychol Sci Soc Sci*. 2021;76(4):692-702
- Inukai, Y., Saito, K., Sasaki, R., Tsuiki, S., Miyaguchi, S., Kojima, S., . . . Onishi, H. (2016). Comparison of Three Non-Invasive Transcranial Electrical Stimulation Methods for Increasing Cortical Excitability. *Front Hum Neurosci*, 10, 668. doi:10.3389/fnhum.2016.00668
- Ironsides M, O'Shea J, Cowen PJ, Harmer CJ. Frontal Cortex Stimulation Reduces Vigilance to Threat: Implications for the Treatment of Depression and Anxiety. *Biol Psychiatry*. 2016 May 15;79(10):823-830.
- Iuculano T, Cohen Kadosh R. The mental cost of cognitive enhancement. *J Neurosci* 2013; 33: 4482-4486
- Ivancovsky, T., Kurman, J., Morio, H., & Shamay-Tsoory, S. (2019). Transcranial direct current stimulation (tDCS) targeting the left inferior frontal gyrus: Effects on creativity across cultures. *Social Neuroscience*, 14(3), 277–285.
- Jackson, M. P., Rahman, A., Lafon, B., Kronberg, G., Ling, D., Parra, L. C., & Bikson, M. (2016). Animal models of transcranial direct current stimulation: Methods and mechanisms. *Clinical Neurophysiology*, 127(11), 3425–3454.
- Jannati, A., Block, G., Oberman, L. M., Rotenberg, A., & Pascual-Leone, A. (2017). Interindividual variability in response to continuous theta-burst stimulation (cTBS) in healthy adults. *Clinical Neurophysiology* : 128(11), 2268–2278.
- Jones AP, Choe J, Bryant NB, Robinson CSH, Ketz NA, Skorheim SW, et al. Dose-Dependent Effects of Closed-Loop tACS Delivered During Slow-Wave Oscillations on Memory Consolidation. *Frontiers in neuroscience*. 2018;12:867.
- Jones, K. T., Stephens, J. A., Alam, M., Bikson, M., & Berryhill, M. E. (2015). Longitudinal Neurostimulation in Older Adults Improves Working Memory. *PLOS ONE*, 10(4), e0121904.
- Juengst, Eric T. 1998. “What Does Enhancement Mean?” In *Enhancing Human Traits: Ethical and Social Implications*, Erik Parens Eds, 29-47. Washington, DC: Georgetown University Press

Jwa, A. 2015. Early adopters of the magical thinking cap: a study on do-it-yourself (DIY) transcranial direct current stimulation (tDCS) user community. *Journal of Law and the Biosciences* 2(2): 292–335. Luculano T, Cohen Kadosh R. The mental cost of cognitive enhancement. *J Neurosci* 2013; 33: 4482-4486

Kadosh RC, Levy N, O’Shea J, Shea N, Savulescu J. The neuroethics of noninvasive brain stimulation. *Curr Biol* 2012; 22: R108e11.

Kadosh, R. C. (2014). *The stimulated brain : cognitive enhancement using non-invasive brain stimulation*. London ; Waltham, MA ; San Diego, CA: Elsevier/Academic Press.

Kan B, Dundas JE, Nosaka K. Effect of transcranial direct current stimulation on elbow flexor maximal voluntary isometric strength and endurance. *Appl Physiol Nutr Metabol* 2013 (38):734-739.

Kasten FH, Dowsett J, Herrmann CS. Sustained Aftereffect of alpha-tACS Lasts Up to 70 min after Stimulation. *Frontiers in human neuroscience*. 2016;10:245.

Kasten FH, Duecker K, Maack MC, Meiser A, Herrmann CS. Integrating electric field modeling and neuroimaging to explain inter-individual variability of tACS effects. *Nature communications*. 2019;10:5427.

Kessler SK, Minhas P, Woods AJ, Rosen A, Gorman C, Bikson M. Dosage considerations for transcranial direct current stimulation in children: a computational modeling study. *PLoS One*. 2013;8:e76112.

Ketz N, Jones AP, Bryant NB, Clark VP, Pilly PK. Closed-Loop Slow-Wave tACS Improves Sleep-Dependent Long-Term Memory Generalization by Modulating Endogenous Oscillations. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2018;38:7314-26.

Kiyonaga A, Egner T. Center-Surround Inhibition in Working Memory. *Current biology : CB*. 2016;26:64-8.

Koo PC, Molle M, Marshall L. Efficacy of slow oscillatory-transcranial direct current stimulation on EEG and memory - contribution of an inter-individual factor. *The European journal of neuroscience*. 2018;47:812-23.

Krause B, Dresler M, Looi CY, Sarkar A, Cohen Kadosh R. Neuroenhancement of High-Level Cognition: Evidence for Homeostatic Constraints of Non-invasive Brain Stimulation. *J Cogn Enhanc*. 2019;3:388-95.

Krause MR, Vieira PG, Csorba BA, Pilly PK, Pack CC (2019) Transcranial alternating current stimulation entrains single-neuron activity in the primate brain. *Proceedings of the National Academy of Sciences*, 116, 5747–5755.

Krause, B., & Cohen, K. R. (2014). Not all brains are created equal: The relevance of individual differences in responsiveness to transcranial electrical stimulation. *Front Syst. Neurosci.*, 8(1662-5137 (Electronic)), 25.

Krebs, C., Peter, J., Wyss, P., Brem, A.-K., & Klöppel, S. (2021). Transcranial electrical stimulation improves cognitive training effects in healthy elderly adults with low cognitive performance. *Clinical Neurophysiology*, 132(6), 1254–1263. <https://doi.org/10.1016/j.clinph.2021.01.034>

Kronberg, G., Rahman, A., Sharma, M., Bikson, M., & Parra, L. C. (2020). Direct current stimulation boosts hebbian plasticity in vitro. *Brain Stimulation*, 13(2), 287–301. <https://doi.org/10.1016/j.brs.2019.10.014>

Kusturica J, Hajdarevic A, Niksic H, Skopljak A, Tafi Z, Kulo A. Neuroenhancing Substances Use, Exam Anxiety and Academic Performance in Bosnian-Herzegovinian First-Year University Students. *Acta Med Acad*. 2019;48:286-93.

Ladenbauer J, Ladenbauer J, Kulzow N, de Boor R, Avramova E, Grittner U, et al. Promoting Sleep Oscillations and Their Functional Coupling by Transcranial Stimulation Enhances Memory Consolidation in Mild Cognitive Impairment. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2017;37:7111-24.

Ladenbauer J, Ladenbauer J, Kulzow N, Floel A. Memory-relevant nap sleep physiology in healthy and pathological aging. *Sleep*. 2021;44.

Lafon B, Henin S, Huang Y, Friedman D, Melloni L, Thesen T, et al. Low frequency transcranial electrical stimulation does not entrain sleep rhythms measured by human intracranial recordings. *Nature communications*. 2017;8:1199.

Lang N, Siebner HR, Ward NS, Lee L, Nitsche MA, Paulus W, et al. How does transcranial DC stimulation of the primary motor cortex alter regional neuronal activity in the human brain? *The European journal of neuroscience*. 2005;22:495-504.

Lattari E, de Oliveira BS, Oliveira BRR, de Mello Pedreiro RC, Machado S, Neto GAM. Effects of transcranial direct current stimulation on time limit and ratings of perceived exertion in physically active women. *Neurosci Lett*. 2018 Jan 1;662:12-16.

- Lattari E, Rosa Filho BJ, Fonseca Junior SJ, Murillo-Rodriguez E, Rocha N, Machado S, Maranhão Neto GA. Effects on Volume Load and Ratings of Perceived Exertion in Individuals' Advanced Weight Training After Transcranial Direct Current Stimulation. *J Strength Cond Res.* 2020 Jan;34(1):89-96.
- Lattari, E., Andrade, M. L., Filho, A. S., Moura, A. M., Neto, G. M., Silva, J. G., et al. (2016). Can transcranial direct current stimulation improve the resistance strength and decrease the rating perceived scale in recreational weight-training experience? *J. Strength Cond. Res.* 30, 3381–3387.
- Lavazza A. The Two-Fold Ethical Challenge in the Use of Neural Electrical Modulation. *Front Neurosci.* 2019 Jun 28;13:678..
- Lee JC, Kenney-Jung DL, Blacker CJ, Doruk Camsari D, Lewis CP. Transcranial Direct Current Stimulation in Child and Adolescent Psychiatric Disorders. *Child Adolesc Psychiatr Clin N Am.* 2019;28:61-78.
- Lee SH, Pitt MA. Individual differences in selective attention reveal the nonmonotonicity of visual spatial attention and its association with working memory capacity. *Journal of experimental psychology General.* 2021.
- Lefaucheur JP, Aleman A, Baeken C, Benninger DH, Brunelin J, Di Lazzaro V, et al. Evidence-based guidelines on the therapeutic use of repetitive transcranial magnetic stimulation (rTMS): An update (2014-2018). *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology.* 2020;131:474-528.
- Lefaucheur JP, Antal A, Ayache SS, Benninger DH, Brunelin J, Cogiamanian F, et al. Evidence-based guidelines on the therapeutic use of transcranial direct current stimulation (tDCS). *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology.* 2017;128:56-92.
- Levasseur-Moreau, J., Brunelin, J., and Fecteau, S. (2013) Noninvasive brain stimulation can induce paradoxical facilitation. Are these neuroenhancements transferable and meaningful to security services? *Frontiers in Human Neuroscience*, 7:449. doi: 10.3389/fnhum.2013.00449
- Liebetanz D, Nitsche MA, Tergau F, Paulus W. Pharmacological approach to the mechanisms of transcranial DC-stimulation-induced after-effects of human motor cortex excitability. *Brain : a journal of neurology.* 2002;125:2238-47.
- Lifshitz-Ben-Basat, A., & Mashal, N. (2021). Enhancing creativity by altering the frontoparietal control network functioning using transcranial direct current stimulation. *Experimental Brain Research*, 239(2), 613–626. <https://doi.org/10.1007/s00221-020-06023-2>
- Looi CY, Duta M, Brem AK, Huber S, Nuerk HC, Cohen Kadosh R. Combining brain stimulation and video game to promote long-term transfer of learning and cognitive enhancement. *Scientific reports.* 2016;6:22003.
- Looi CY, Lim J, Sella F, Lolliot S, Duta M, Avramenko AA, et al. Transcranial random noise stimulation and cognitive training to improve learning and cognition of the atypically developing brain: A pilot study. *Scientific reports.* 2017;7:4633.
- Lorenz R, Simmons LE, Monti RP, Arthur JL, Limal S, Laakso I, et al. Efficiently searching through large tACS parameter spaces using closed-loop Bayesian optimization. *Brain Stimul* 2019;12:1484–9.
- Luber, B., Stanford, A.D., Bulow, P., Nguyen, T., Rakitin, B.C., Habeck, C., Basner, R., Stern, Y. and Lisanby, S.H. (2008) Remediation of sleep-deprivation induced visual working memory impairment with fMRI-guided Transcranial Magnetic Stimulation. *Cerebral Cortex* 18 (9): 2077-2085.
- Luber, B., and Lisanby, S. H. (2014) Enhancement of human cognitive performance using transcranial magnetic stimulation (TMS). *NeuroImage* 85, 961–70.
- Luber, B., Kinnunen, L.H., Rakitin, B.C., Ellsasser, R., Stern, Y., and Lisanby, S.H. (2007) Facilitation of performance in a working memory task with rTMS stimulation of the precuneus: Frequency and time-dependent effects. *Brain Research*, 1128: 120-129.
- Luber, B., Steffener, J., Tucker, A., Habeck, C., Peterchev, A.P., Deng, Z.-D., Basner, R.C., Stern, Y., & Lisanby, S.H. (2013) Extended remediation of sleep deprived-induced working memory deficits using fMRI-guided transcranial magnetic stimulation. *Sleep*, 36(6): 857-871.
- Lucchiari, C., Sala, P. M., & Vanutelli, M. E. (2018). Promoting Creativity Through Transcranial Direct Current Stimulation (tDCS). A Critical Review. *Frontiers in Behavioral Neuroscience*, 12, 167. <https://doi.org/10.3389/fnbeh.2018.00167>

- Lustenberger C, Boyle MR, Alagapan S, Mellin JM, Vaughn BV, Frohlich F. Feedback-Controlled Transcranial Alternating Current Stimulation Reveals a Functional Role of Sleep Spindles in Motor Memory Consolidation. *Current biology* : CB. 2016;26:2127-36.
- Lustenberger, C., Boyle, M. R., Foulser, A. A., Mellin, J. M., & Fröhlich, F. (2015). Functional role of frontal alpha oscillations in creativity. *Cortex; a Journal Devoted to the Study of the Nervous System and Behavior*, 67, 74–82.
- Machado S, Jansen P, Almeida V, Veldema J. Is tDCS an Adjunct Ergogenic Resource for Improving Muscular Strength and Endurance Performance? A Systematic Review. *Front Psychol*. 2019 May 16;10:1127.
- Malish, R.G. (2017) The importance of the study of cognitive enhancement for U.S. national security. *Aerospace Medicine and Human Performance*, 88: 773-778.
- Malkani RG, Zee PC. Brain Stimulation for Improving Sleep and Memory. *Sleep Med Clin*. 2020;15:101-15.
- Mancuso, L. E., Ilieva, I. P., Hamilton, R. H., & Farah, M. J. (2016). Does Transcranial Direct Current Stimulation Improve Healthy Working Memory?: A Meta-analytic Review. *Journal of Cognitive Neuroscience*, 28(8), 1063–1089.
- Manor, B., Zhou, J., Jor'dan, A., Zhang, J., Fang, J., & Pascual-Leone, A. (2016). Reduction of Dual-task Costs by Noninvasive Modulation of Prefrontal Activity in Healthy Elders. *Journal of Cognitive Neuroscience*, 28(2), 275–281.
- Marshall L, Helgadottir H, Molle M, Born J. Boosting slow oscillations during sleep potentiates memory. *Nature*. 2006;444:610-3.
- Martin, D. M., Liu, R., Alonzo, A., Green, M., & Loo, C. K. (2014). Use of transcranial direct current stimulation (tDCS) to enhance cognitive training: effect of timing of stimulation. *Experimental Brain Research*, 232(10), 3345-3351.
- Maslen H, Douglas T, Cohen Kadosh R, Levy N, Savulescu J. The regulation of cognitive enhancement devices: extending the medical model. *Journal of law and the biosciences*. 2014;1:68-93.
- Matsumoto H, Ugawa Y. Quadripulse stimulation (QPS). *Exp Brain Res* 238(7), 1619-1625, 2020
- Matsumoto H, Ugawa Y. Adverse events of tDCS and tACS: A review. *Clin Neurophysiol* 2017;2:19–25.
- Mauger, A. R. (2013). Fatigue is a pain-the use of novel neurophysiological techniques to understand the fatigue-pain relationship. *Front. Physiol*. 4:104.
- Mayseless, N., & Shamay-Tsoory, S. G. (2015). Enhancing verbal creativity: Modulating creativity by altering the balance between right and left inferior frontal gyrus with tDCS. *Neuroscience*, 291, 167–176.
- McCall, I. C., C. Lau, N. Minielly, and J. Illes. 2019. Owing Ethical Innovation: Claims about Commercial Wearable Brain Technologies. *Neuron* 102(4): 728–731.
- McCombs, M. E., & Shaw, D. L. (1972). The agenda-setting function of mass media. *Public Opinion Quarterly*, 36, 176-187.
- McCombs, M. E., Shaw, D. L., & Weaver, D. H. (2014). New directions in agenda-setting theory and research. *Mass Communication and Society*, 17(6), 781-802.
- McCormick, A., Meijen, C., and Marcora, S. (2015). Psychological determinants of whole-body endurance performance. *Sport Med*. 45, 997–1015.
- McIntire LK, McKinley RA, Nelson JM, Goodyear C. Transcranial direct current stimulation versus caffeine as a fatigue countermeasure. *Brain stimulation*. 2017;10:1070-8.
- McKinley, R. A., Bridges, N., Walters, C. M., & Nelson, J. (2012). Modulating the brain at work using noninvasive transcranial stimulation. *NeuroImage*, 59, 129–137.
- McKinley, R. A., McIntire, L., Bridges, N., Goodyear, C., Bangera, N.B., and Weisend, M.P. (2013) Acceleration of image analyst training with transcranial direct current stimulation. *Behavioral Neuroscience*, 127: 936-946.
- Menotti F, Berchicci M, Di Russo F, Damiani A, Vitelli S, Macaluso A. The role of the prefrontal cortex in the development of muscle fatigue in Charcot-Marie-Tooth 1A patients. *Neuromuscul Disord*. 2014 Jun;24(6):516-23.

Moliadze V, Andreas S, Lyzhko E, Schmanke T, Gurashvili T, Freitag CM, et al. Ten minutes of 1 mA transcranial direct current stimulation was well tolerated by children and adolescents: Self-reports and resting state EEG analysis. *Brain research bulletin*. 2015a;119:25-33.

Moliadze V, Atalay D, Antal A, Paulus W. Close to threshold transcranial electrical stimulation preferentially activates inhibitory networks before switching to excitation with higher intensities. *Brain Stimul*. 2012 Oct;5(4):505-11

Moliadze V, Lyzhko E, Schmanke T, Andreas S, Freitag CM, Siniatchkin M. 1mA cathodal tDCS shows excitatory effects in children and adolescents: Insights from TMS evoked N100 potential. *Brain research bulletin*. 2018;140:43-51.

Moliadze V, Schmanke T, Andreas S, Lyzhko E, Freitag CM, Siniatchkin M. Stimulation intensities of transcranial direct current stimulation have to be adjusted in children and adolescents. *Clinical neurophysiology* .: 2015b;126:1392-9.

Monte-Silva K, Kuo MF, Hessenthaler S, Fresnoza S, Liebetanz D, Paulus W, et al. Induction of late LTP-like plasticity in the human motor cortex by repeated non-invasive brain stimulation. *Brain stimulation*. 2013;6:424-32.

Morgan HM, Davis NJ, Bracewell RM. Does transcranial direct current stimulation to prefrontal cortex affect mood and emotional memory retrieval in healthy individuals? *PloS one*. 2014;9:e92162.

Mosayebi Samani M, Agboada D, Jamil A, Kuo MF, Nitsche MA. Titrating the neuroplastic effects of cathodal transcranial direct current stimulation (tDCS) over the primary motor cortex. *Cortex; a journal devoted to the study of the nervous system and behavior*. 2019;119:350-61.

Mosayebi Samani M, Agboada D, Kuo MF, Nitsche MA. Probing the relevance of repeated cathodal transcranial direct current stimulation over the primary motor cortex for prolongation of after-effects. *The Journal of physiology*. 2020;598:805-16.

Motohashi N, Yamaguchi M, Fujii T, Kitahara Y. Mood and cognitive function following repeated transcranial direct current stimulation in healthy volunteers: a preliminary report. *Neuroscience research*. 2013;77:64-9.

Mulquiney, P. G., Hoy, K. E., Daskalakis, Z. J., & Fitzgerald, P. B. (2011). Improving working memory: Exploring the effect of transcranial random noise stimulation and transcranial direct current stimulation on the dorsolateral prefrontal cortex. *Clinical Neurophysiology*, 122(12), 2384-2389.

Murphy, O. W., Hoy, K. E., Wong, D., Bailey, N. W., Fitzgerald, P. B., & Segrave, R. A. (2020). Transcranial random noise stimulation is more effective than transcranial direct current stimulation for enhancing working memory in healthy individuals: Behavioural and electrophysiological evidence. *Brain Stimul*, 13(5), 1370-1380. doi:10.1016/j.brs.2020.07.001

Muszkat D, Polanczyk GV, Dias TG, Brunoni AR. Transcranial Direct Current Stimulation in Child and Adolescent Psychiatry. *J Child Adolesc Psychopharmacol*. 2016;26:590-7.

Muthalib M, Kan B, Nosaka K, Perrey S. Effects of transcranial direct current stimulation of the motor cortex on prefrontal cortex activation during a neuromuscular fatigue task: an fNIRS study. *Adv Exp Med Biol* 2013 (789):73.

Neri F, Mencarelli L, Menardi A, Giovannelli F, Rossi S, Sprugnoli G, et al. A novel tDCS sham approach based on model-driven controlled shunting. *Brain Stimul* 2020;13:507–16.

Neubauer, A. C., Wammerl, M., Benedek, M., Jauk, E., & Jaušovec, N. (2017). The influence of transcranial alternating current stimulation (tACS) on fluid intelligence: An fMRI study. *Personality and Individual Differences*, 118, 50–55.

Nikolin, S., Martin, D., Loo, C. K., & Boonstra, T. W. (2018). Effects of TDCS dosage on working memory in healthy participants. *Brain Stimul*, 11(3), 518-527.

Nissim, N. R., Moberg, P. J., & Hamilton, R. H. (2020). Efficacy of Noninvasive Brain Stimulation (tDCS or TMS) Paired with Language Therapy in the Treatment of Primary Progressive Aphasia: An Exploratory Meta-Analysis. *Brain Sciences*, 10(9), 597.

Nitsche M, Schauenburg A, Lang N, Liebetanz D, Exner C, Paulus W, et al. Facilitation of implicit motor learning by weak transcranial direct current stimulation of the primary motor cortex in the human. *Journal of cognitive neuroscience*. 2003;15:619-26.

Nitsche MA, Doemkes S, Karaköse T, Antal A, Liebetanz D, Lang N, Tergau F, Paulus W. Shaping the effects of transcranial direct current stimulation of the human motor cortex. *J Neurophysiol.* 2007;97(4):3109-17

Noakes, T. D. (2012). Fatigue is a brain-derived emotion that regulates the exercise behavior to ensure the protection of whole body homeostasis. *Front. Physiol.* 3, 1–13.

O’Connell NE, Cossar J, Marston L, Wand BM, Bunce D, Moseley GL, et al. Rethinking clinical trials of transcranial direct current stimulation: participant and assessor blinding is inadequate at intensities of 2mA. *PLoS One* 2012;7:e47514.

Oberman, L.M., Exley, S., Philip, N.S., Siddiqi, S.H., Adamson, M.M., and Brody, D.L. (2020) The use of transcranial magnetic stimulation in the treatment of neuropsychiatric and neurocognitive symptoms associated with concussion in military populations. *Journal of Head Trauma and Rehabilitation*, 35: 388-400.

Okano, A. H., Fontes, E. B., Montenegro, R. A., Farinatti, Pde, T., Cyrino, E. S., Noakes, T. D., et al. (2015). Brain stimulation modulates the autonomic nervous system, rating of perceived exertion and performance during maximal exercise. *Br. J. Sports Med.* 49, 1213–1218.

Oken BS, Flegal K, Zajdel D, Kishiyama S, Haas M, Peters D. Expectancy effect: Impact of pill administration on cognitive performance in healthy seniors. *J Clin Exp Neuropsychol* 2008;30:7–17..

Ozdemir, R. A., Tadayon, E., Boucher, P., Sun, H., Momi, D., Ganglberger, W., Westover, M. B., Pascual-Leone, A., Santarnecchi, E., & Shafi, M. M. (2021). Cortical responses to noninvasive perturbations enable individual brain fingerprinting. *Brain Stimulation*, 14(2), 391–403.

Pahor, A., & Jaušovec, N. (2014). The effects of theta transcranial alternating current stimulation (tACS) on fluid intelligence. *International Journal of Psychophysiology: Official Journal of the International Organization of Psychophysiology*, 93(3), 322–331.

Palm U, Reisinger E, Keeser D, Kuo M-F, Pogarell O, Leicht G, et al. Evaluation of sham transcranial direct current stimulation for randomized, placebo-controlled clinical trials. *Brain Stimul* 2013;6:690–5.

Palm U, Segmiller FM, Epple AN, Freisleder FJ, Koutsouleris N, Schulte-Korne G, et al. Transcranial direct current stimulation in children and adolescents: a comprehensive review. *Journal of neural transmission (Vienna, Austria : 1996)*. 2016;123:1219-34.

Parasuraman, R. and Galster, C. (2013) Sensing, assessing, and augmenting threat detection: Behavioral, neuroimaging, and brain stimulation evidence for the critical role of attention. *Frontiers in Human Neuroscience*, 7, Article 273.

Park K. Neuro-doping: the rise of another loophole to get around anti-doping policies. *Cogent Soc Sci* 2017 (3):4-11.

Paulus W, Rothwell JC. Membrane resistance and shunting inhibition: where biophysics meets state-dependent human neurophysiology. *The Journal of physiology*. 2016;594:2719-28.

Paus, T., Sipila, P. K., and Strafella, A. P. (2001b). Synchronization of neuronal activity in the human primary motor cortex by transcranial magnetic stimulation: an EEG study. *J. Neurophysiol.* 86, 1983–1990.

Peña, C. T. de la. 2003. *The Body Electric: How Strange Machines Built the Modern American*. New York University Press. New York University Press.

Peña, J., Sampedro, A., Ibarretxe-Bilbao, N., Zubiaurre-Elorza, L., & Ojeda, N. (2019). Improvement in creativity after transcranial random noise stimulation (trNS) over the left dorsolateral prefrontal cortex. *Scientific Reports*, 9(1), 7116.

Petersen TH, Puthusserypady S. Assessing tDCS Placebo Effects on EEG and Cognitive Tasks. *Annu Int Conf IEEE Eng Med Biol Soc IEEE Eng Med Biol Soc* 2019:4509–12.

Petrocchi N, Piccirillo G, Fiorucci C, Moscucci F, Di Iorio C, Mastropietri F, et al. Transcranial direct current stimulation enhances soothing positive affect and vagal tone. *Neuropsychologia*. 2017;96:256-61.

Plewnia C, Schroeder PA, Kunze R, Faehling F, Wolkenstein L. Keep calm and carry on: improved frustration tolerance and processing speed by transcranial direct current stimulation (tDCS). *PLoS one*. 2015;10:e0122578.

Prehn K, Flöel A. Potentials and limits to enhance cognitive functions in healthy and pathological aging by tDCS. *Front Cell Neurosci*. 2015 Sep 14;9:355.

Prehn-Kristensen A, Munz M, Goder R, Wilhelm I, Korr K, Vahl W, et al. Transcranial oscillatory direct current stimulation during sleep improves declarative memory consolidation in children with attention-deficit/hyperactivity disorder to a level comparable to healthy controls. *Brain stimulation*. 2014;7:793-9.

Pugh J, Pugh C. Neurostimulation, doping, and the spirit of sport. *Neuroethics*. 2021;14(Suppl 2):141-158.

Racine E, Nguyen V, Saigle V, Dubljevic V. Media Portrayal of a Landmark Neuroscience Experiment on Free Will. *Science and engineering ethics*. 2017;23:989-1007.

Racine E, Waldman S, Palmour N, Risse D, Illes J. "Currents of hope": neurostimulation techniques in U.S. and U.K. print media. *Camb Q Healthc Ethics*. 2007;16:312-6.

Rajapakse T, Kirton A. NON-INVASIVE BRAIN STIMULATION IN CHILDREN: APPLICATIONS AND FUTURE DIRECTIONS. *Transl Neurosci*. 2013;4:10.2478/s13380-013-0116-3.

Rangarajan SK, Suhas S, Reddy MSS, Sreeraj VS, Sivakumar PT, Venkatasubramanian G. Domiciliary tDCS in Geriatric Psychiatric Disorders: Opportunities and Challenges. *Indian Journal of Psychological Medicine* 2021 (online ahead)

Rasch B, Born J. About sleep's role in memory. *Physiol Rev*. 2013;93:681-766.

Reato, D., Rahman, A., Bikson, M., & Parra, L. C. (2013). Effects of weak transcranial alternating current stimulation on brain activity—A review of known mechanisms from animal studies. *Frontiers in Human Neuroscience*, 7.

Reis J, Robertson EM, Krakauer JW, Rothwell J, Marshall L, Gerloff C, et al. Consensus: Can transcranial direct current stimulation and transcranial magnetic stimulation enhance motor learning and memory formation? *Brain stimulation*. 2008;1:363-9.

Reis J, Schambra HM, Cohen LG, Buch ER, Fritsch B, Zarahn E, et al. Noninvasive cortical stimulation enhances motor skill acquisition over multiple days through an effect on consolidation. *Proceedings of the National Academy of Sciences of the United States of America*. 2009;106:1590-5.

Repantis, D., Bovy, L., Ohla, K., Kuhn, S., & Dresler, M. (2021). Cognitive enhancement effects of stimulants: a randomized controlled trial testing methylphenidate, modafinil, and caffeine. *Psychopharmacology (Berl)*, 238(2), 441-451

Rivera-Urbina GN, Nitsche MA, Vicario CM, Molero-Chamizo A. Applications of transcranial direct current stimulation in children and pediatrics. *Rev Neurosci*. 2017;28:173-84.

Robinson CSH, Bryant NB, Maxwell JW, Jones AP, Robert B, Lamphere M, et al. The Benefits of Closed-Loop Transcranial Alternating Current Stimulation on Subjective Sleep Quality. *Brain sciences*. 2018;8.

Robinson, C., Armenta, M., Combs, A., Lamphere, M. L., Garza, G. J., Neary, J., Wolfe, J. H., Molina, E., Semey, D. E., McKee, C. M., Gallegos, S. J., Jones, A. P., Trumbo, M. C., Al-Azzawi, H., Hunter, M. A., Lieberman, G., Coffman, B. A., Abozeria, M., Bikson, M., ... Witkiewitz, K. (2019). Modulating affective experience and emotional intelligence with loving kindness meditation and transcranial direct current stimulation: A pilot study. *Social Neuroscience*, 14(1), 10–25.

Rosanova, M., Casali, A., Bellina, V., Resta, F., Mariotti, M., and Massimini, M. (2009). natural frequencies of human corticothalamic circuits. *J. Neurosci*. 29, 7679–7685.

Rossi S, Antal A, Bestmann S, Bikson M, Brewer C, Brockmoller J, et al. Safety and recommendations for TMS use in healthy subjects and patient populations, with updates on training, ethical and regulatory issues: Expert Guidelines. *Clinical neurophysiology* : 2021;132:269-306.

Rossini PM, Di Iorio R, Bentivoglio M, Bertini G, Ferreri F, Gerloff C, Ilmoniemi RJ, Miraglia F, Nitsche MA, Pestilli F, Rosanova M, Shirota Y, Tesoriero C, Ugawa Y, Vecchio F, Ziemann U, Hallett M. Methods for analysis of brain connectivity: An IFCN-sponsored review. *Clin Neurophysiol*. 2019 Oct;130(10):1833-1858.

Ruffini, G., Modolo, J., Sanchez-Todo, R., Salvador, R., & Santarnecchi, E. (2020). Clinical Drivers for Personalization of Transcranial Current Stimulation (tES 3.0). *Non Invasive Brain Stimulation in Psychiatry and Clinical Neurosciences*, 353–370.

Sahlem GL, Badran BW, Halford JJ, Williams NR, Korte JE, Leslie K, et al. Oscillating Square Wave Transcranial Direct Current Stimulation (tDCS) Delivered During Slow Wave Sleep Does Not Improve Declarative Memory More Than Sham: A Randomized Sham Controlled Crossover Study. *Brain stimulation*. 2015;8:528-34.

- Sakai K, Hikosaka O, Miyauchi S, Takino R, Sasaki Y, Pütz B. Transition of brain activation from frontal to parietal areas in visuo-motor sequence learning. *J Neurosci* 1998; 18: 1827-1840.
- Sale, M. V., Mattingley, J. B., Zaleski, A., and Cocchi, L. (2015). Imaging human brain networks to improve the clinical efficacy of noninvasive brain stimulation. *Neurosci. Biobehav. Rev.* 57, 187–198.
- Salehinejad MA, Wischniewski M, Nejati V, Vicario CM, Nitsche MA. Transcranial direct current stimulation in attention-deficit hyperactivity disorder: A meta-analysis of neuropsychological deficits. *PLoS One*. 2019;14:e0215095-e.
- Sales, M. M., De Sousa, C. V., Browne, R. A. V., Fontes, E. B., Olher, R. R. V., Ernesto, C., et al. (2016). Transcranial direct current stimulation improves muscle isokinetic performance of young trained individuals. *Med. Dello Sport* 69, 1–10.
- Sanchez-Leon, C. A., Sanchez-Lopez, A., Gomez-Climent, M. A., Cordones, I., Cohen Kadosh, R., & Marquez-Ruiz, J. (in press). Impact of chronic transcranial Random Noise Stimulation (tRNS) on GABAergic and glutamatergic activity markers in the prefrontal cortex of juvenile mice. *Progress in Brain Research*. doi:10.1101/2020.09.04.282889
- Santarnecchi E, Feurra M, Galli G, Rossi A, Rossi S. Overclock your brain for gaming? Ethical, social and health care risks. *Brain stimulation*. 2013;6:713-4.
- Santarnecchi, E., Muller, T., Rossi, S., Sarkar, A., Polizzotto, N. R., Rossi, A., & Cohen Kadosh, R. (2016). Individual differences and specificity of prefrontal gamma frequency-tACS on fluid intelligence capabilities. *Cortex*, 75, 33-43.
- Santarnecchi, E., Polizzotto, N. R., Godone, M., Giovannelli, F., Feurra, M., Matzen, L., Rossi, A., & Rossi, S. (2013). Frequency-dependent enhancement of fluid intelligence induced by transcranial oscillatory potentials. *Current Biology: CB*, 23(15), 1449–1453.
- Sarkar A, Dowker A, Cohen Kadosh R. Cognitive enhancement or cognitive cost: trait-specific outcomes of brain stimulation in the case of mathematics anxiety. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2014;34:16605-10.
- Sarkar, A., Dowker, A., & Cohen Kadosh, R. (2014). Cognitive Enhancement or Cognitive Cost: Trait-Specific Outcomes of Brain Stimulation in the Case of Mathematics Anxiety. *Journal of Neuroscience*, 34(50), 16605–16610. <https://doi.org/10.1523/JNEUROSCI.3129-14.2014>
- Saturnino GB, Antunes A, Thielscher A. On the importance of electrode parameters for shaping electric field patterns generated by tDCS. *Neuroimage*. 2015;120:25-35.
- Schecklmann, M., Nejati, V., Poepl, T. B., Peytard, J., Rupprecht, R., Wetter, T. C., Kreuzer, P. M. (2021). Bifrontal high-frequency transcranial random noise stimulation is not effective as an add-on treatment in depression. *Journal of Psychiatric Research*, 132, 116-122.
- Schubert, M. M., and Astorino, T. A. (2012). A systematic review of the efficacy of ergogenic aids for improving running performance. *J. Strength Cond. Res.* 27, 1699–1707.
- Schutter DJLG (2016). Cutaneous retinal activation and neural entrainment in transcranial alternating current stimulation: A systematic review. *NeuroImage*, 140, 83–88.
- Schutter DJLG, Hortensius R. Retinal origin of phosphenes to transcranial alternating current stimulation. *Clin Neurophysiol* 2010;121:1080–4.
- Sehm, B. and Ragert, P. (2013) Why non-invasive brain stimulation should not be used in military and security services. *Frontiers in Human Neuroscience*, 7: 553.
- Sellers KK, Mellin JM, Lustenberger CM, Boyle MR, Lee WH, Peterchev AV, et al. Transcranial direct current stimulation (tDCS) of frontal cortex decreases performance on the WAIS-IV intelligence test. *Behavioural brain research*. 2015;290:32-44.
- Sheffield, J. G., Raz, G., Sella, F., & Cohen Kadosh, R. (2020). How can noise alter neurophysiology in order to improve human behaviour? A combined tRNS and EEG study. *bioRxiv*, 2020.2001.2009.900118.
- Shimizu T, Hanajima R, Shirota Y, Tsutsumi R, Tanaka N, Terao Y, Hamada M, Ugawa Y. (2020) Plasticity induction in the pre-supplementary motor area (pre-SMA) and SMA-proper differentially affects visuomotor sequence learning. *Brain Stimul* 13:229-238.

- Shirota Y, Hewitt M, Paulus W. Neuroscientists do not use non-invasive brain stimulation on themselves for neural enhancement. *Brain stimulation*. 2014;7:618-9.
- Shorafa Y, Halawa I, Hewitt M, Nitsche MA, Antal A, Paulus W. Isometric agonist and antagonist muscle activation interacts differently with 140-Hz transcranial alternating current stimulation aftereffects at different intensities. *Journal of neurophysiology*. 2021;126:340-8.
- Siddiqi, S. H., Weigand, A., Pascual-Leone, A., & Fox, M. D. (2021). Identification of Personalized Transcranial Magnetic Stimulation Targets Based on Subgenual Cingulate Connectivity: An Independent Replication. *Biological Psychiatry*, 0(0).
- Sierawska, A., Moliadze, V., Splittgerber, M., Rogge, A., Siniatchkin, M., & Buyx, A. (2020). First Epileptic Seizure and Initial Diagnosis of Juvenile Myoclonus Epilepsy (JME) in a Transcranial Direct Current Stimulation (tDCS) Study- Ethical Analysis of a Clinical case. *Neuroethics*, 13(3), 347-351.
- Silvanto, J., Pascual-Leone, A., 2008. State-dependency of transcranial magnetic stimulation. *Brain Topogr*. 21, 1–10.
- Snowball, A., Tachtsidis, I., Popescu, T., Thompson, J., Delazer, M., Zamarian, L., Cohen Kadosh, R. (2013). Long-Term Enhancement of Brain Function and Cognition Using Cognitive Training and Brain Stimulation. *Current Biology*, 23(11), 987-992.
- Splittgerber M, Suwelack JH, Kadish NE, Moliadze V. The Effects of 1 mA tACS and tRNS on Children/Adolescents and Adults: Investigating Age and Sensitivity to Sham Stimulation. *Neural plasticity*. 2020;2020:8896423.
- Splittgerber, M., Borzikowsky, C., Salvador, R., Puonti, O., Papadimitriou, K., Merschformann, C., . . . Moliadze, V. (2021). Multichannel anodal tDCS over the left dorsolateral prefrontal cortex in a paediatric population. *Sci Rep*, 11(1), 21512.
- Splittgerber, M., Japaridze, N., Sierawska, A., Gimenez, S., Nowak, R., Siniatchkin, M., & Moliadze, V. (2020). First generalized tonic clonic seizure in the context of pediatric tDCS - A case report. *Neurophysiol Clin*, 50(1), 69-72.
- Steenbergen L, Sellaro R, Hommel B, Lindenberger U, Kühn S, Colzato LS. "Unfocus" on foc.us: commercial tDCS headset impairs working memory. *Exp Brain Res*. 2016, 234: 637-43
- Stein, M. (n.d.). Creativity and culture. *Journal of Psychology*, 36, 31–322.
- Stewart B, Outram S, & Smith AC. Doing supplements to improve performance in club cycling: A life-course analysis. *Scandinavian Journal of Medicine & Science in Sports* 2013 (23):361–372.
- Strenze, T. (2007). Intelligence and socioeconomic success: A meta-analytic review of longitudinal research. *Intelligence*, 35(5), 401–426.
- Tanaka M, Ishii A, Watanabe Y. Neural mechanism of facilitation system during physical fatigue. *PLoS One*. 2013 Nov 20;8(11):e80731.
- Terney D, Chaieb L, Moliadze V, Antal A, Paulus W. Increasing human brain excitability by transcranial high-frequency random noise stimulation. *The Journal of neuroscience : the official journal of the Society for Neuroscience*. 2008;28:14147-55.
- Thut, G., Schyns, P. G., Gross, J., 2011. Entrainment of perceptually relevant brain oscillations by non-invasive rhythmic stimulation of the human brain. *Front. Psychol.* 2, 170.
- Tiksnadi A, Murakami T, Wiratman W, Matsumoto H, Ugawa Y. (2020) Direct comparison of efficacy of the motor cortical plasticity induction and their inter-individual variability between TBS and QPS. *Brain Stimul* 13:1824-1833.
- Tononi G, Cirelli C. Sleep and synaptic down-selection. *The European journal of neuroscience*. 2020;51:413-21.
- Turi Z, Ambrus GG, Janacsek K, Emmert K, Hahn L, Paulus W, et al. Both the cutaneous sensation and phosphene perception are modulated in a frequency-specific manner during transcranial alternating current stimulation. *Restor Neurol Neurosci* 2013;31:275–85.
- Turi Z, Bjørkedal E, Gunkel L, Antal A, Paulus W, Mittner M. Evidence for Cognitive Placebo and Nocebo Effects in Healthy Individuals. *Sci Rep* 2018;8:17443.
- Turi Z, Csifcsák G, Boayue NM, Aslaksen P, Antal A, Paulus W, et al. Blinding is compromised for transcranial direct current stimulation at 1 mA for 20 min in young healthy adults. *Eur J Neurosci* 2019;50.

- Turi Z, Mittner M, Paulus W, Antal A. Placebo Intervention Enhances Reward Learning in Healthy Individuals. *Sci Rep* 2017;7:41028.
- Turkeltaub, P. E., Benson, J., Hamilton, R. H., Datta, A., Bikson, M., & Coslett, H. B. (2012). Left lateralizing transcranial direct current stimulation improves reading efficiency. *Brain Stimulation*, 5(3), 201–207.
- Turner C, Jackson C, Learmonth G. Is the “end-of-study guess” a valid measure of sham blinding during transcranial direct current stimulation? *Eur J Neurosci* 2020.
- van Bueren NER, Reed TL, Nguyen V, Sheffield JG, van der Ven SHG, Osborne MA, et al. Personalized brain stimulation for effective neurointervention across participants. *PLoS computational biology*. 2021;17:e1008886.
- Van Cutsem, J., De Pauw, K., Buyse, L., Marcora, S., Meeusen, R., and Roelands, B. (2017a). Effects of mental fatigue on endurance performance in the heat. *Med. Sci. Sports Exerc.* 49, 1677–1687.
- Van Noppen P, van Dun K, Depestele S, Verstraelen S, Meesen R, Manto M. Transcranial direct current stimulation and attention skills in burnout patients: a randomized blinded sham-controlled pilot study. *F1000Research*. 2020;9:116.
- Vargas, V. Z., Baptista, A. F., Pereira, G. O. C., Pochini, A. C., Ejnisman, B., Santos, M. B., et al. (2018). Modulation of isometric quadriceps strength in soccer players with transcranial direct current stimulation: a crossover study. *J. Strength Cond. Res.* 32, 1336–1341.
- Vaseghi, B., Zoghi, M., and Jaberzadeh, S. (2014). Does anodal transcranial direct current stimulation modulate sensory perception and pain? A meta-analysis study. *Clin. Neurophysiol.* 125, 1847–1858.
- Vieira LAF, Lattari E, de Jesus Abreu MA, Rodrigues GM, Viana B, Machado S, Oliveira BRR, Maranhão Neto GA. Transcranial Direct Current Stimulation (tDCS) Improves Back-Squat Performance in Intermediate Resistance-Training Men. *Res Q Exerc Sport*. 2020 Sep 15:1-9.
- Vieira PG, Krause MR, Pack CC (2020) tACS entrains neural activity while somatosensory input is blocked. *PLoS Biology*, 18, e3000834.
- Voarino N, Dubljevic V, Racine E. tDCS for Memory Enhancement: Analysis of the Speculative Aspects of Ethical Issues. *Frontiers in human neuroscience*. 2016;10:678.
- Vöröslakos M, Takeuchi Y, Brinyiczki K, Zombori T, Oliva A, Fernández-Ruiz A, Kozák G, Kincses ZT, Iványi B, Buzsáki G, Berényi A (2018) Direct effects of transcranial electric stimulation on brain circuits in rats and humans. *Nature Communications*, 9, 483.
- Voskuhl J, Struber D, Herrmann CS. Non-invasive Brain Stimulation: A Paradigm Shift in Understanding Brain Oscillations. *Frontiers in human neuroscience*. 2018;12:211.
- Wager TD, Atlas LY. The neuroscience of placebo effects: Connecting context, learning and health. *Nat Rev Neurosci* 2015;16:403–18.
- Waits, R. K. 2013. *The Medical Electricians: George A. Scott and His Victorian Cohorts in Quackery*. CreateSpace Independent Publishing Platform. CreateSpace Independent Publishing Platform.
- Waltz, E. 2019. The brain hackers. *Nature Biotechnology* 37(9): 1–5. doi: 10.1038/s41587-019-0238-4.
- Wang L, Yu H, Hu J, Theeuwes J, Gong X, Xiang Y, et al. Reward breaks through center-surround inhibition via anterior insula. *Human brain mapping*. 2015;36:5233-51.
- Wang, B., Xiao, S., Yu, C., Zhou, J., & Fu, W. (2021). Effects of Transcranial Direct Current Stimulation Combined With Physical Training on the Excitability of the Motor Cortex, Physical Performance, and Motor Learning: A Systematic Review. *Frontiers in Neuroscience*, 15, 648354.
- Weller, S., Nitsche, M. A., & Plewnia, C. (2020). Enhancing cognitive control training with transcranial direct current stimulation: A systematic parameter study. *Brain Stimulation*, 13(5), 1358–1369.
- Wexler, A. 2015. A pragmatic analysis of the regulation of consumer transcranial direct current stimulation (tDCS) devices in the United States. *Journal of Law and the Biosciences* 2(3): 669–696.
- Wexler, A. 2016. The practices of do-it-yourself brain stimulation: implications for ethical considerations and regulatory proposals. *Journal of Medical Ethics* 42(4): 211–215. doi: 10.1136/medethics-2015-102704.

- Wexler, A. 2017a. Recurrent themes in the history of the home use of electrical stimulation: Transcranial direct current stimulation (tDCS) and the medical battery (1870–1920). *Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation* 10(2): 187–195.
- Wexler, A. 2017b. The Medical Battery in The United States (1870-1920): Electrotherapy at Home and in the Clinic. *Journal of the history of medicine and allied sciences* 72(2): 166–192.
- Wexler, A. 2018. Who Uses Direct-to-Consumer Brain Stimulation Products, and Why? A Study of Home Users of tDCS Devices. *Journal of Cognitive Enhancement* 2(1): 114–134.
- Wexler, A. 2020. Do-it-yourself and direct-to-consumer neurostimulation. *Developments in Neuroethics and Bioethics*. doi: 10.1016/bs.dnb.2020.03.005.
- Woods AJ, Antal A, Bikson M, Boggio PS, Brunoni AR, Celnik P, Cohen LG, Fregni F, Herrmann CS, Kappenman ES, Knotkova H, Liebetanz D, Miniussi C, Miranda PC, Paulus W, Priori A, Reato D, Stagg C, Wenderoth N, Nitsche MA. A technical guide to tDCS, and related non-invasive brain stimulation tools. *Clinical Neurophysiology*. 2016; 127(2):1031-48.
- Wurzman R, Hamilton RH, Pascual-Leone A, Fox MD. An open letter concerning do-it-yourself users of transcranial direct current stimulation. *Ann Neurol* 2016, 80(1): 1-4
- Zrenner, B., Zrenner, C., Gordon, P. C., Belardinelli, P., McDermott, E. J., Soekadar, S. R., Fallgatter, A. J., Ziemann, U., & Müller-Dahlhaus, F. (2020). Brain oscillation-synchronized stimulation of the left dorsolateral prefrontal cortex in depression using real-time EEG-triggered TMS. *Brain Stimulation*, 13(1), 197–205.
- Zrenner, C., Desideri, D., Belardinelli, P., and Ziemann, U. (2018) Real-time EEG-defined excitability states determine efficacy of TMS-induced plasticity in human motor cortex. *Brain Stimulation*, 11: 374-389.
- Zulkifly MFM, Merkohitaj O, Brockmoller J, Paulus W. Confounding effects of caffeine on neuroplasticity induced by transcranial alternating current stimulation and paired associative stimulation. *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology*. 2021a;132:1367-79.
- Zulkifly MFM, Merkohitaj O, Paulus W, Brockmoller J. The roles of caffeine and corticosteroids in modulating cortical excitability after paired associative stimulation (PAS) and transcranial alternating current stimulation (tACS) in caffeine-naive and caffeine-adapted subjects. *Psychoneuroendocrinology*. 2021b;127:105201.
- Zulkifly MFM, Merkohitaj O, Paulus W. Transcranial alternating current stimulation induced excitatory aftereffects are abolished by decaffeinated espresso and reversed into inhibition by espresso with caffeine. *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology*. 2020;131:778-9.